



### Analytical Summary Report

Analysis Method: 9045D  
Parameter: pH  
Run Number: LB129390  
BalanceID: WC SC-4

Analyst By : jignesh  
Supervisor Review By : Iwona  
Slope : 98.5  
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/05/2024	14:50
2	CAL2	1	Water	NA	NA	20.3	6.99	02/05/2024	14:51
3	CAL3	1	Water	NA	NA	20.3	10.03	02/05/2024	14:52
4	ICV	1	Water	NA	NA	20.2	7.00	02/05/2024	14:55
5	CCV1	1	Water	NA	NA	20.3	2.01	02/05/2024	14:59
6	P1374-01	1	Solid	20.02	20	22.1	7.39	02/05/2024	15:05
7	P1374-01DUP	1	Solid	20.03	20	22.2	7.40	02/05/2024	15:06
8	CCV2	1	Water	NA	NA	20.3	12.02	02/05/2024	15:10

UB 129390

### WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH P1374      WorkList ID : 177646      Department : Wet-Chemistry      Date : 02-05-2024 14:35:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1374-01	BU-03-020524	Solid	pH	Cool 4 deg C	PSEG05	D31	02/05/2024	9045D

Date/Time 02/05/24 14:40  
 Raw Sample Received by: SO CWF  
 Raw Sample Relinquished by: OP SR

Date/Time 02/05/24 17:00  
 Raw Sample Received by: OP SR  
 Raw Sample Relinquished by: SO CWF