



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB129393

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/05/2024	14:50
2	CAL2	1	Water	NA	NA	20.3	6.99	02/05/2024	14:51
3	CAL3	1	Water	NA	NA	20.3	10.03	02/05/2024	14:52
4	ICV	1	Water	NA	NA	20.2	7.00	02/05/2024	14:55
5	CCV1	1	Water	NA	NA	20.4	2.02	02/05/2024	15:15
6	P1373-05	1	Water	NA	NA	20.4	6.39	02/05/2024	15:20
7	P1373-05DUP	1	Water	NA	NA	20.5	6.40	02/05/2024	15:21
8	CCV2	1	Water	NA	NA	20.3	12.02	02/05/2024	15:25

WORKLIST(Hardcopy Internal Chain)

W3 129393

WorkList Name : PH P1373

WorkList ID : 177647

Department : Wet-Chemistry

Date : 02-05-2024 14:35:56

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1373-05	HT-020524	Water	pH	Cool 4 deg C	PSEG03	D31	02/05/2024	9040C

Date/Time 02/05/24 14:40
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 02/05/24 17:10C
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]