

LB129395


245.1

INSTRUMENT ID: CV4

Linear

 μ Abs.:

41170



A= 0.0000e+000

B= 2.4559e-004

C= -5.8457e-002

Rho= 0.9998992

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%p
0.0	0.000	-0.020	-0.020	155	0.000	155					-18
0.05	0.050	0.041	-0.009	405	0.0 %	405	0				-2
0.2	0.200	0.197	-0.003	1039	0.0 %	1039					4
2.5	2.500	2.589	0.089	10778	0.0 %	10778					0
5.0	5.000	4.981	-0.019	20519	0.0 %	20519					-1
7.5	7.500	7.411	-0.089	30413	0.0 %	30413					-1
10.0	10.000	10.053	0.053	41170	0.0 %	41170					

LB129395 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 05 Feb 2024 15:31:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	80	155	-	0.0000	245.1	PPB	05 Feb 2024 16:01:35	S	Std
0.05 - 1	50.05	405	-	0.0500	245.1	PPB	05 Feb 2024 16:06:33	S	Std
0.2 - 1	50.2	1039	-	0.2000	245.1	PPB	05 Feb 2024 16:11:49	S	Std
2.5 - 1	52.5	10778	-	2.5000	245.1	PPB	05 Feb 2024 16:14:07	S	Std
5.0 - 1	45	20519	-	5.0000	245.1	PPB	05 Feb 2024 16:16:23	S	Std
7.5 - 1	67.5	30413	-	7.5000	245.1	PPB	05 Feb 2024 16:18:41	S	Std
10.0 - 1	510	41170	-	10.0000	245.1	PPB	05 Feb 2024 16:24:03	S	Std
ICV95 - 1	ICV95	16409	3.9715	-245.1		PPB	05 Feb 2024 16:26:53	U	SMPL
ICB95 - 1	ICB95	141	-0.0238	-245.1		PPB	05 Feb 2024 16:31:55	U	SMPL
CCV72 - 1	CCV72	20312	4.9300	-245.1		PPB	05 Feb 2024 16:34:10	U	SMPL
CCB72 - 1	CCB72	37	-0.0494	-245.1		PPB	05 Feb 2024 16:41:50	U	SMPL
CRA - 1	CRA	408	0.0417	-245.1		PPB	05 Feb 2024 16:44:05	U	SMPL
HighStd - 1	HighStd	41598	10.1577	-245.1		PPB	05 Feb 2024 16:46:20	U	SMPL
ChkStd - 1	ChkStd	27784	6.7651	-245.1		PPB	05 Feb 2024 16:48:35	U	SMPL
PB158789BL - 1	PBW	85	-0.0376	-245.1		PPB	05 Feb 2024 16:53:30	U	SMPL
PB158789BS - 1	LCSW	16992	4.1146	-245.1		PPB	05 Feb 2024 16:55:45	U	SMPL
O4699-10 - 1		385	0.0361	-245.1		PPB	05 Feb 2024 17:14:43	U	SMPL
O4699-10RE - 1		848	0.1498	-245.1		PPB	05 Feb 2024 17:17:00	U	SMPL
O4699-12 - 1	MDL-WATER-06-QT4-2023	423	0.0454	-245.1		PPB	05 Feb 2024 17:26:49	U	SMPL
CCV73 - 1	CCV73	20728	5.0322	-245.1		PPB	05 Feb 2024 17:31:22	U	SMPL
CCB73 - 1	CCB73	124	-0.0280	-245.1		PPB	05 Feb 2024 17:36:18	U	SMPL
P1214-01 - 1	EFF-DAY-1-METALS	882	0.1582	-245.1		PPB	05 Feb 2024 17:38:33	U	SMPL
P1214-01DUP - 1	EFF-DAY-1-METALSDUP	634	0.0972	-245.1		PPB	05 Feb 2024 17:40:49	U	SMPL
P1214-01MS - 1	EFF-DAY-1-METALSMS	17078	4.1358	-245.1		PPB	05 Feb 2024 17:43:03	U	SMPL
P1214-01MSD - 1	EFF-DAY-1-METALSMSD	16426	3.9756	-245.1		PPB	05 Feb 2024 17:45:19	U	SMPL
P1327-01 - 1	MANHOLE	-129	-0.0901	-245.1		PPB	05 Feb 2024 17:47:37	U	SMPL
P1174-05 - 1	PT-HG-WS	21690	5.2684	-245.1		PPB	05 Feb 2024 17:54:30	U	SMPL
P1214-01LX5 - 1		296	0.0142	-245.1		PPB	05 Feb 2024 17:56:50	U	SMPL
P1214-01A - 1		16793	4.0658	-245.1		PPB	05 Feb 2024 17:59:10	U	SMPL
CCV74 - 1	CCV74	19457	4.7200	-245.1		PPB	05 Feb 2024 18:01:25	U	SMPL
CCB74 - 1	CCB74	-59	-0.0729	-245.1		PPB	05 Feb 2024 18:03:43	U	SMPL