


LB129402

7470A

INSTRUMENT ID: CV1

 μ Abs.:

45352



Linear

A= 0.0000e+000

B= 2.1798e-004

C= -1.0238e-001

Rho= 0.9990844

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/10
0.0	0.000	-0.051	-0.051	235	0.000	235	0				
0.2	0.200	0.180	-0.020	1295	0.0 %	1295					10
2.5	2.500	2.566	0.066	12243	0.0 %	12243					3
5.0	5.000	4.923	-0.077	23053	0.0 %	23053					-2
7.5	7.500	7.799	0.299	36247	0.0 %	36247					4
10.0	10.000	9.783	-0.217	45352	0.0 %	45352					2

LB129402 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Feb 2024 10:12:39

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	80	235	-	0.0000	7470A	PPB	06 Feb 2024 10:20:26	S	Std
0.2 - 1	50.2	1295	-	0.2000	7470A	PPB	06 Feb 2024 10:22:43	S	Std
2.5 - 1	52.5	12243	-	2.5000	7470A	PPB	06 Feb 2024 10:25:00	S	Std
5.0 - 1	55	23053	-	5.0000	7470A	PPB	06 Feb 2024 10:27:17	S	Std
7.5 - 1	57.5	36247	-	7.5000	7470A	PPB	06 Feb 2024 10:29:34	S	Std
10.0 - 1	510	45352	-	10.0000	7470A	PPB	06 Feb 2024 10:31:53	S	Std
ICV96 - 1	ICV96	18141	3.8520	-	7470A	PPB	06 Feb 2024 10:34:48	U	SMPL
ICB96 - 1	ICB96	343	-0.0276	-	7470A	PPB	06 Feb 2024 10:37:03	U	SMPL
CCV75 - 1	CCV75	23342	4.9857	-	7470A	PPB	06 Feb 2024 10:42:01	U	SMPL
CCB75 - 1	CCB75	433	-0.0080	-	7470A	PPB	06 Feb 2024 10:44:16	U	SMPL
CRA - 1	CRA	1218	0.1631	-	7470A	PPB	06 Feb 2024 10:46:33	U	SMPL
HighStd - 1	HighStd	44279	9.5496	-	7470A	PPB	06 Feb 2024 10:48:49	U	SMPL
ChkStd - 1	ChkStd	32343	6.9477	-	7470A	PPB	06 Feb 2024 10:51:05	U	SMPL
PB158830BL - 1	PBW	360	-0.0239	-	7470A	PPB	06 Feb 2024 10:56:01	U	SMPL
PB158830BS - 1	LCSW	19592	4.1683	-	7470A	PPB	06 Feb 2024 10:58:18	U	SMPL
P1066-02 - 1	WC-N-COMP	-213	-0.1488	-	7470A	PPB	06 Feb 2024 11:00:34	U	SMPL
P1141-02 - 1	EB-3-0-2	16	-0.0989	-	7470A	PPB	06 Feb 2024 11:02:54	U	SMPL
P1141-04 - 1	EB-4-0-2	76	-0.0858	-	7470A	PPB	06 Feb 2024 11:05:10	U	SMPL
P1141-06 - 1	EB-3-2-4	258	-0.0461	-	7470A	PPB	06 Feb 2024 11:07:27	U	SMPL
P1141-08 - 1	EB-4-6-8	291	-0.0389	-	7470A	PPB	06 Feb 2024 11:09:42	U	SMPL
CCV76 - 1	CCV76	24361	5.2078	-	7470A	PPB	06 Feb 2024 11:11:58	U	SMPL
CCB76 - 1	CCB76	494	0.0053	-	7470A	PPB	06 Feb 2024 11:14:13	U	SMPL
P1141-10 - 1	EB-2-0-2	380	-0.0195	-	7470A	PPB	06 Feb 2024 11:16:31	U	SMPL
P1141-12 - 1	EB-2-0-2D	447	-0.0049	-	7470A	PPB	06 Feb 2024 11:18:46	U	SMPL
P1318-04 - 1	CO-11-GRID-1	386	-0.0182	-	7470A	PPB	06 Feb 2024 11:21:01	U	SMPL
P1318-05 - 1	CO-11-GRID-2	348	-0.0265	-	7470A	PPB	06 Feb 2024 11:23:17	U	SMPL
P1336-03 - 1	356	226	-0.0531	-	7470A	PPB	06 Feb 2024 11:25:33	U	SMPL
P1336-06 - 1	RT4008	244	-0.0492	-	7470A	PPB	06 Feb 2024 11:27:49	U	SMPL
P1337-03 - 1	TP-1	211	-0.0564	-	7470A	PPB	06 Feb 2024 11:30:05	U	SMPL
P1337-06 - 1	TP-2	153	-0.0690	-	7470A	PPB	06 Feb 2024 11:32:23	U	SMPL
P1337-09 - 1	TP-3	130	-0.0740	-	7470A	PPB	06 Feb 2024 11:34:39	U	SMPL
P1337-12 - 1	TP-4	162	-0.0671	-	7470A	PPB	06 Feb 2024 11:36:55	U	SMPL
CCV77 - 1	CCV77	21814	4.6526	-	7470A	PPB	06 Feb 2024 11:46:29	U	SMPL
CCB77 - 1	CCB77	408	-0.0134	-	7470A	PPB	06 Feb 2024 11:48:44	U	SMPL
P1337-15 - 1	TP-5	256	-0.0466	-	7470A	PPB	06 Feb 2024 11:51:02	U	SMPL
P1337-19 - 1	TP-6	156	-0.0684	-	7470A	PPB	06 Feb 2024 11:53:17	U	SMPL
P1338-02 - 1	WC-S-COMP	132	-0.0736	-	7470A	PPB	06 Feb 2024 11:55:33	U	SMPL
P1338-02DUP - 1	WC-S-COMPDUP	118	-0.0767	-	7470A	PPB	06 Feb 2024 11:57:49	U	SMPL
P1338-02MS - 1	WC-S-COMPMS	18148	3.8535	-	7470A	PPB	06 Feb 2024 12:02:59	U	SMPL
P1338-02MSD - 1	WC-S-COMPMSD	17567	3.7269	-	7470A	PPB	06 Feb 2024 12:05:15	U	SMPL
PB158831BL - 1	PBW	95	-0.0817	-	7470A	PPB	06 Feb 2024 12:07:33	U	SMPL
PB158831BS - 1	LCSW	17732	3.7628	-	7470A	PPB	06 Feb 2024 12:12:34	U	SMPL
P1355-03 - 1	VNJ-214	132	-0.0736	-	7470A	PPB	06 Feb 2024 12:14:50	U	SMPL
P1360-01 - 1	D3366	54	-0.0906	-	7470A	PPB	06 Feb 2024 12:17:08	U	SMPL
CCV78 - 1	CCV78	22629	4.8303	-	7470A	PPB	06 Feb 2024 12:19:24	U	SMPL
CCB78 - 1	CCB78	461	-0.0019	-	7470A	PPB	06 Feb 2024 12:21:39	U	SMPL
P1360-02 - 1	D3367	25	-0.0969	-	7470A	PPB	06 Feb 2024 12:23:58	U	SMPL
P1360-03 - 1	D3368	462	-0.0017	-	7470A	PPB	06 Feb 2024 12:26:13	U	SMPL
P1360-05 - 1	D3370	364	-0.0230	-	7470A	PPB	06 Feb 2024 12:28:29	U	SMPL
P1362-04 - 1	WC-1	286	-0.0400	-	7470A	PPB	06 Feb 2024 12:30:44	U	SMPL
P1362-08 - 1	WC-2	113	-0.0777	-	7470A	PPB	06 Feb 2024 12:33:02	U	SMPL
P1362-12 - 1	WC-3	92	-0.0823	-	7470A	PPB	06 Feb 2024 12:35:19	U	SMPL
P1366-02 - 1	EB-2-6-8	414	-0.0121	-	7470A	PPB	06 Feb 2024 12:37:37	U	SMPL
P1366-04 - 1	EB-1-0-2	67	-0.0878	-	7470A	PPB	06 Feb 2024 12:39:55	U	SMPL
P1366-06 - 1	EB-1-4-6	79	-0.0852	-	7470A	PPB	06 Feb 2024 12:42:13	U	SMPL
P1367-02 - 1	EB-6-0-2	111	-0.0782	-	7470A	PPB	06 Feb 2024 12:44:28	U	SMPL
CCV79 - 1	CCV79	22154	4.7267	-	7470A	PPB	06 Feb 2024 12:46:44	U	SMPL
CCB79 - 1	CCB79	484	0.0031	-	7470A	PPB	06 Feb 2024 12:49:04	U	SMPL
P1367-04 - 1	EB-7-0-2	35	-0.0948	-	7470A	PPB	06 Feb 2024 12:51:26	U	SMPL
P1367-06 - 1	EB-6-7-9	75	-0.0860	-	7470A	PPB	06 Feb 2024 12:53:42	U	SMPL
P1367-08 - 1	EB-7-5-7	146	-0.0706	-	7470A	PPB	06 Feb 2024 12:55:58	U	SMPL
P1367-10 - 1	EB-7-5-7D	121	-0.0760	-	7470A	PPB	06 Feb 2024 12:58:13	U	SMPL
P1367-12 - 1	EB-5-0-2	126	-0.0749	-	7470A	PPB	06 Feb 2024 13:00:29	U	SMPL
P1367-14 - 1	EB-5-6-8	108	-0.0788	-	7470A	PPB	06 Feb 2024 13:02:45	U	SMPL
P1367-14DUP - 1	EB-5-6-8DUP	89	-0.0830	-	7470A	PPB	06 Feb 2024 13:05:02	U	SMPL
P1367-14MS - 1	EB-5-6-8MS	15824	3.3469	-	7470A	PPB	06 Feb 2024 13:07:19	U	SMPL
P1367-14MSD - 1	EB-5-6-8MSD	14584	3.0766	-	7470A	PPB	06 Feb 2024 13:09:35	U	SMPL
PB158848BL - 1	PBW	307	-0.0355	-	7470A	PPB	06 Feb 2024 13:14:42	U	SMPL
CCV80 - 1	CCV80	24345	5.2043	-	7470A	PPB	06 Feb 2024 13:19:42	U	SMPL
CCB80 - 1	CCB80	129	-0.0743	-	7470A	PPB	06 Feb 2024 13:21:57	U	SMPL
PB158848BS - 1	LCSW	16023	3.3903	-	7470A	PPB	06 Feb 2024 13:24:16	U	SMPL
P1342-06 - 1	MR-323P-GW-20240131-0	639	0.0369	-	7470A	PPB	06 Feb 2024 13:26:32	U	SMPL
P1342-07 - 1	MR-323P-GW-20240131-1	668	0.0432	-	7470A	PPB	06 Feb 2024 13:28:50	U	SMPL
P1342-08 - 1	FB-20240131	-78	-0.1194	-	7470A	PPB	06 Feb 2024 13:31:08	U	SMPL
P1342-09 - 1	MR-317AP-GW-20240131-0	269	-0.0437	-	7470A	PPB	06 Feb 2024 13:33:25	U	SMPL
P1354-01 - 1	FRAC-TANK-A6014	3751	0.7153	-	7470A	PPB	06 Feb 2024 13:35:41	U	SMPL
P1365-03 - 1	FB-20240201	52	-0.0910	-	7470A	PPB	06 Feb 2024 13:37:58	U	SMPL
P1368-01 - 1	MR-306P-GW-20240201-0	14592	3.0784	-	7470A	PPB	06 Feb 2024 13:40:16	U	SMPL
P1371-06 - 1	T385-AOP	298	-0.0374	-	7470A	PPB	06 Feb 2024 13:42:34	U	SMPL
P1372-01 - 1	CG-MW-63	151	-0.0695	-	7470A	PPB	06 Feb 2024 13:44:52	U	SMPL
CCV81 - 1	CCV81	24026	5.1348	-	7470A	PPB	06 Feb 2024 13:47:10	U	SMPL

LB129402 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Feb 2024 10:12:39

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB81 - 1	CCB81	258	-0.0461		-7470A	PPB	06 Feb 2024 13:49:26	U	SMPL
P1372-01DUP - 1	CG-MW-63DUP	128	-0.0745		-7470A	PPB	06 Feb 2024 13:51:44	U	SMPL
P1372-02 - 1	P1372-01MS	18185	3.8616		-7470A	PPB	06 Feb 2024 13:56:49	U	SMPL
P1372-03 - 1	P1372-01MSD	17969	3.8145		-7470A	PPB	06 Feb 2024 13:59:04	U	SMPL
PB158779TB - 1	PB158779TB	248	-0.0483		-7470A	PPB	06 Feb 2024 14:04:09	U	SMPL
PB158806TB - 1	PB158806TB	203	-0.0581		-7470A	PPB	06 Feb 2024 14:06:25	U	SMPL
CCV82 - 1	CCV82	21820	4.6539		-7470A	PPB	06 Feb 2024 14:08:41	U	SMPL
CCB82 - 1	CCB82	195	-0.0599		-7470A	PPB	06 Feb 2024 14:13:44	U	SMPL
P1338-02LX5 - 1		-86	-0.1211		-7470A	PPB	06 Feb 2024 14:16:00	U	SMPL
P1338-02A - 1		19069	4.0543		-7470A	PPB	06 Feb 2024 14:33:38	U	SMPL
P1367-14LX5 - 1		-298	-0.1673		-7470A	PPB	06 Feb 2024 14:35:54	U	SMPL
P1367-14A - 1		17364	3.6826		-7470A	PPB	06 Feb 2024 14:38:12	U	SMPL
P1372-01LX5 - 1		-277	-0.1628		-7470A	PPB	06 Feb 2024 14:40:29	U	SMPL
P1372-01A - 1		19366	4.1190		-7470A	PPB	06 Feb 2024 14:42:48	U	SMPL
CCV83 - 1	CCV83	24426	5.2220		-7470A	PPB	06 Feb 2024 14:45:04	U	SMPL
CCB83 - 1	CCB83	97	-0.0812		-7470A	PPB	06 Feb 2024 14:47:23	U	SMPL