



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129408

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	02/06/2024	08:35
2	P1294-01		1	12	100.00	>212.0	>212.0	02/06/2024	09:05
3	P1294-01DUP		1	12	100.00	>212.0	>212.0	02/06/2024	09:35
4	P1304-03		1	14	100.00	>212.0	>212.0	02/06/2024	10:05
5	P1304-04		1	13	100.00	>212.0	>212.0	02/06/2024	10:35
6	P1312-01		1	14	100.00	>212.0	>212.0	02/06/2024	11:05
7	P1312-02		1	12	100.00	>212.0	>212.0	02/06/2024	11:35
8	P1354-01		1	14	100.00	>212.0	>212.0	02/06/2024	12:05
9	P1357-01		1	14	100.00	>212.0	>212.0	02/06/2024	12:35
10	P1357-02		1	12	100.00	>212.0	>212.0	02/06/2024	13:05
11	P1360-04		1	13	100.00	>212.0	>212.0	02/06/2024	13:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-02-05

WorkList ID : 177671

Department : Wet-Chemistry

Date : 02-05-2024 08:19:05

6129408

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1294-01	MMW-001	Water	Flash Point	Cool 4 deg C	ENTA05	D11	01/30/2024	1010B
P1304-03	BRIGHTON-MANHOLE-1-8TH-	Water	Flash Point	Cool 4 deg C	TULL02	D11	01/30/2024	1010B
P1304-04	BRIGHTON-MANHOLE-2-8TH-	Water	Flash Point	Cool 4 deg C	TULL02	D11	01/30/2024	1010B
P1312-01	NWB-1999	Water	Flash Point	Cool 4 deg C	PSEG03	E41	01/31/2024	1010B
P1312-02	NWW-2005	Water	Flash Point	Cool 4 deg C	PSEG03	E41	01/31/2024	1010B
P1354-01	FRAC-TANK-A6014	Water	Flash Point	Cool 4 deg C	PSEG03	D11	02/02/2024	1010B
P1357-01	50264	Water	Flash Point	Cool 4 deg C	PSEG03	D11	02/02/2024	1010B
P1357-02	50261	Water	Flash Point	Cool 4 deg C	PSEG03	D11	02/02/2024	1010B
P1360-04	D3369	Water	Flash Point	Cool 4 deg C	PSEG03	E11	02/02/2024	1010B

Date/Time 02/06/2024 08:25
 Raw Sample Received by: RMY
 Raw Sample Relinquished by: MGL

Date/Time 02/06/2024 14:15
 Raw Sample Received by: MGL
 Raw Sample Relinquished by: RMY