


LB129409

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.7587e-004

C= -3.6988e-002

Rho= 0.9999321

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.015	0.015	298	0.000	298					1/10
0.2	0.200	0.226	0.026	1493	0.0 %	1493					13
2.5	2.500	2.444	-0.056	14108	0.0 %	14108					-2
5.0	5.000	4.968	-0.032	28460	0.0 %	28460					-1
7.5	7.500	7.573	0.073	43272	0.0 %	43272					-1
10.0	10.000	9.973	-0.027	56918	0.0 %	56918					0

LB129409 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Feb 2024 15:09:26

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	298	-	0.0000	7471B	PPB	06 Feb 2024 15:15:16	S	Std
0.2 - 1	502	1493	-	0.2000	7471B	PPB	06 Feb 2024 15:17:33	S	Std
2.5 - 1	52.5	14108	-	2.5000	7471B	PPB	06 Feb 2024 15:19:50	S	Std
5.0 - 1	55	28460	-	5.0000	7471B	PPB	06 Feb 2024 15:22:06	S	Std
7.5 - 1	57.5	43272	-	7.5000	7471B	PPB	06 Feb 2024 15:27:09	S	Std
10.0 - 1	510	56918	-	10.0000	7471B	PPB	06 Feb 2024 15:29:28	S	Std
ICV97 - 1	ICV97	24044	4.1916	-	7471B	PPB	06 Feb 2024 15:32:21	U	SMPL
ICB97 - 1	ICB97	478	0.0471	-	7471B	PPB	06 Feb 2024 15:34:38	U	SMPL
CCV84 - 1	CCV84	28566	4.9869	-	7471B	PPB	06 Feb 2024 15:37:02	U	SMPL
CCB84 - 1	CCB84	501	0.0511	-	7471B	PPB	06 Feb 2024 15:39:17	U	SMPL
CRA - 1	CRA	1450	0.2180	-	7471B	PPB	06 Feb 2024 15:41:44	U	SMPL
HighStd - 1	HighStd	57160	10.0158	-	7471B	PPB	06 Feb 2024 15:44:00	U	SMPL
ChkStd - 1	ChkStd	43859	7.6765	-	7471B	PPB	06 Feb 2024 15:50:47	U	SMPL
PB158852BL - 1	PBS	122	-0.0155	-	7471B	PPB	06 Feb 2024 15:53:03	U	SMPL
PB158852BS - 1	LCSS	20337	3.5397	-	7471B	PPB	06 Feb 2024 15:55:32	U	SMPL
P1066-01 - 1	WC-N-COMP	10239	1.7638	-	7471B	PPB	06 Feb 2024 15:57:48	U	SMPL
P1337-01 - 1	TP-1	1546	0.2349	-	7471B	PPB	06 Feb 2024 16:00:13	U	SMPL
P1337-04 - 1	TP-2	1448	0.2177	-	7471B	PPB	06 Feb 2024 16:02:33	U	SMPL
P1337-07 - 1	TP-3	1167	0.1683	-	7471B	PPB	06 Feb 2024 16:04:52	U	SMPL
P1337-10 - 1	TP-4	2892	0.4716	-	7471B	PPB	06 Feb 2024 16:07:07	U	SMPL
CCV85 - 1	CCV85	28908	5.0471	-	7471B	PPB	06 Feb 2024 16:09:22	U	SMPL
CCB85 - 1	CCB85	462	0.0443	-	7471B	PPB	06 Feb 2024 16:11:39	U	SMPL
P1337-13 - 1	TP-5	3222	0.5297	-	7471B	PPB	06 Feb 2024 16:14:04	U	SMPL
P1337-17 - 1	TP-6	2102	0.3327	-	7471B	PPB	06 Feb 2024 16:21:23	U	SMPL
P1337-17DUP - 1	TP-6DUP	2124	0.3366	-	7471B	PPB	06 Feb 2024 16:23:38	U	SMPL
P1337-17MS - 1	TP-6MS	23540	4.1030	-	7471B	PPB	06 Feb 2024 16:25:55	U	SMPL
P1337-17MSD - 1	TP-6MSD	24380	4.2507	-	7471B	PPB	06 Feb 2024 16:28:12	U	SMPL
P1338-01 - 1	WC-S-COMP	3838	0.6380	-	7471B	PPB	06 Feb 2024 16:30:37	U	SMPL
P1345-01 - 1	TT-278-IDWSO-20240201-1	867	0.1155	-	7471B	PPB	06 Feb 2024 16:33:02	U	SMPL
P1355-01 - 1	VNJ-214	2509	0.4043	-	7471B	PPB	06 Feb 2024 16:35:22	U	SMPL
P1356-01 - 1	OR-3-020224	3234	0.5318	-	7471B	PPB	06 Feb 2024 16:37:40	U	SMPL
P1358-01 - 1	RT2302	1508	0.2282	-	7471B	PPB	06 Feb 2024 16:39:57	U	SMPL
CCV86 - 1	CCV86	29207	5.0997	-	7471B	PPB	06 Feb 2024 16:42:14	U	SMPL
CCB86 - 1	CCB86	531	0.0564	-	7471B	PPB	06 Feb 2024 16:44:29	U	SMPL
P1358-03 - 1	CHRT24449	7380	1.2609	-	7471B	PPB	06 Feb 2024 16:46:58	U	SMPL
P1358-05 - 1	RT5149	1680	0.2585	-	7471B	PPB	06 Feb 2024 16:49:13	U	SMPL
P1337-17LX5 - 1		436	0.0397	-	7471B	PPB	06 Feb 2024 16:51:31	U	SMPL
P1337-17A - 1		24838	4.3313	-	7471B	PPB	06 Feb 2024 16:53:47	U	SMPL
CCV87 - 1	CCV87	29402	5.1340	-	7471B	PPB	06 Feb 2024 16:56:02	U	SMPL
CCB87 - 1	CCB87	331	0.0212	-	7471B	PPB	06 Feb 2024 16:58:26	U	SMPL