

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

3/5/2024 13:15

Sample id	Test id	Result	Status	Date	Errors
ICV1	Total CN	93.595	man acc	2/6/2024	
ICB1	Total CN	0.963	man acc	2/6/2024	
CCV1	Total CN	246.976	man acc	2/6/2024	
CCB1	Total CN	1.026	man acc	2/6/2024	
RL CHECK	Total CN	5.133	man acc	2/6/2024	102% (50-150)
PB158842BL 6684	Total CN	1.159	man acc	2/6/2024	
PB158842BL 6685	Total CN	95.114	man acc	2/6/2024	02/06/24
MIDPB158842 66 12	Total CN	248.402	man acc	2/6/2024	89% (90-110) 12
P1327-01	Total CN	19.120	man acc	2/6/2024	
P1327-01DUP	Total CN	19.294	man acc	2/6/2024	
P1327-01MS	Total CN	58.457	man acc	2/6/2024	
P1327-01MSD	Total CN	59.570	man acc	2/6/2024	
P1368 9-01 12	Total CN	1.696	man acc	2/6/2024	
CCV2	Total CN	254.718	man acc	2/6/2024	
CCB2	Total CN	2.101	man acc	2/6/2024	

Aquakem v. 7.2AQ1

Results from time period:

Tue Feb 06 13:44:43 2024

Tue Feb 06 14:30:46 2024

Sample Id	Sam/Ctr/c/A	Test short n	Test type	Result	Result unit	Result date and time Stat
0.0PPBCN	A	Total CN	P	1.178	µg/l	2/6/2024 9:36:30
5.0PPBCN	A	Total CN	P	5.2818	µg/l	2/6/2024 9:36:31
10PPBCN	A	Total CN	P	9.9163	µg/l	2/6/2024 9:36:32
50PPBCN	A	Total CN	P	47.4464	µg/l	2/6/2024 9:36:33
100PPBCN	A	Total CN	P	97.6326	µg/l	2/6/2024 9:36:34
250PPBCN	A	Total CN	P	255.6342	µg/l	2/6/2024 9:36:35
500PPBCN	A	Total CN	P	497.9106	µg/l	2/6/2024 9:36:36
ICV1	S	Total CN	P	93.5948	µg/l	2/6/2024 13:44:44
ICB1	S	Total CN	P	0.9632	µg/l	2/6/2024 13:44:45
CCV1	S	Total CN	P	246.9761	µg/l	2/6/2024 13:44:47
CCB1	S	Total CN	P	1.0257	µg/l	2/6/2024 13:44:50
RL CHECK	S	Total CN	P	5.1332	µg/l	2/6/2024 13:44:51
PB158866BL	S	Total CN	P	1.1589	µg/l	2/6/2024 13:52:14
PB158866BS	S	Total CN	P	95.114	µg/l	2/6/2024 13:52:16
MIDPB158866	S	Total CN	P	248.4015	µg/l	2/6/2024 13:52:17
P1327-01	S	Total CN	P	19.1203	µg/l	2/6/2024 13:52:20
P1327-01DUP	S	Total CN	P	19.294	µg/l	2/6/2024 13:52:22
P1327-01MS	S	Total CN	P	58.4568	µg/l	2/6/2024 13:59:23
P1327-01MSD	S	Total CN	P	59.57	µg/l	2/6/2024 13:59:26
P1369-01	S	Total CN	P	1.6959	µg/l	2/6/2024 13:59:27
CCV2	S	Total CN	P	254.7176	µg/l	2/6/2024 13:59:29
CCB2	S	Total CN	P	2.1014	µg/l	2/6/2024 13:59:30

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

2/6/2024 9:36

Reviewed by : RM

Instrument ID : Konelab


Test Total CN

Accepted 2/6/2024 9:36

Factor 1256
Bias 0.001

Coeff. of det. 0.999758

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.002	1.1780	0.0000	-
2	5.0PPBCN	0.005	5.2818	5.0000	5.6
3	10PPBCN	0.008	9.9163	10.0000	-6.8
4	50PPBCN	0.038	47.4464	50.0000	-5.1
5	100PPBCN	0.078	97.6326	100.0000	-2.4
6	250PPBCN	0.204	255.6342	250.0000	2.3
7	500PPBCN	0.397	497.9106	500.0000	-0.4

02/06/2024
RM