

LB129420


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

5094E



A= 0.0000e+000

B= 1.9391e-004

C= -1.1664e-001

Rho= 0.9991134

Accept=Accepted

Max. Conc: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
0.0	0.000	-0.069	-0.069	247	0.000	247					
0.2	0.200	0.200	0.000	1633	0.0 %	1633					
2.5	2.500	2.509	0.009	13539	0.0 %	13539					
5.0	5.000	5.009	0.009	26432	0.0 %	26432					
7.5	7.500	7.789	0.289	40768	0.0 %	40768					
10.0	10.000	9.762	-0.238	50945	0.0 %	50945					

10
0
0
0
4
-2

LB129420 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 07 Feb 2024 11:17:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	247	-	0.0000	7471B	PPB	07 Feb 2024 11:32:14	S	Std
0.2 - 1	50.2	1633	-	0.2000	7471B	PPB	07 Feb 2024 11:34:31	S	Std
2.5 - 1	12.5	13539	-	2.5000	7471B	PPB	07 Feb 2024 11:36:48	S	Std
5.0 - 1	55	26432	-	5.0000	7471B	PPB	07 Feb 2024 11:39:05	S	Std
7.5 - 1	57.5	40768	-	7.5000	7471B	PPB	07 Feb 2024 11:41:25	S	Std
10.0 - 1	57.0	50945	-	10.0000	7471B	PPB	07 Feb 2024 11:43:52	S	Std
ICV98 - 1	ICV98	22991	4.3416	-	7471B	PPB	07 Feb 2024 11:49:25	U	SMPL
ICB98 - 1	ICB98	762	0.0311	-	7471B	PPB	07 Feb 2024 11:51:44	U	SMPL
CCV88 - 1	CCV88	24419	4.6185	-	7471B	PPB	07 Feb 2024 11:54:09	U	SMPL
CCB88 - 1	CCB88	711	0.0212	-	7471B	PPB	07 Feb 2024 11:56:23	U	SMPL
CRA - 1	CRA	1585	0.1907	-	7471B	PPB	07 Feb 2024 11:58:51	U	SMPL
HighStd - 1	HighStd	52051	9.9767	-	7471B	PPB	07 Feb 2024 12:01:06	U	SMPL
ChkStd - 1	ChkStd	38669	7.3818	-	7471B	PPB	07 Feb 2024 12:03:21	U	SMPL
PB158873BL - 1	PBS	89	-0.0994	-	7471B	PPB	07 Feb 2024 12:05:57	U	SMPL
PB158873BS - 1	LCSS	19440	3.6530	-	7471B	PPB	07 Feb 2024 12:08:25	U	SMPL
P1141-01 - 1	EB-3-0-2	12117	2.2330	-	7471B	PPB	07 Feb 2024 12:10:42	U	SMPL
P1141-03 - 1	EB-4-0-2	14003	2.5987	-	7471B	PPB	07 Feb 2024 12:13:05	U	SMPL
P1141-05 - 1	EB-3-2-4	16046	2.9949	-	7471B	PPB	07 Feb 2024 12:15:25	U	SMPL
P1141-07 - 1	EB-4-6-8	2541	0.3761	-	7471B	PPB	07 Feb 2024 12:17:47	U	SMPL
P1141-09 - 1	EB-2-0-2	8241	1.4814	-	7471B	PPB	07 Feb 2024 12:20:06	U	SMPL
CCV89 - 1	CCV89	24545	4.6430	-	7471B	PPB	07 Feb 2024 12:22:23	U	SMPL
CCB89 - 1	CCB89	768	0.0323	-	7471B	PPB	07 Feb 2024 12:24:42	U	SMPL
P1141-11 - 1	EB-2-0-2D	11285	2.0717	-	7471B	PPB	07 Feb 2024 12:27:10	U	SMPL
P1378-01 - 1	TP-4	28179	5.3476	-	7471B	PPB	07 Feb 2024 12:29:25	U	SMPL
P1378-03 - 1	TP-5	2657	0.3986	-	7471B	PPB	07 Feb 2024 12:31:44	U	SMPL
P1390-01 - 1	ETGI-357	1121	0.1007	-	7471B	PPB	07 Feb 2024 12:34:09	U	SMPL
P1391-01 - 1	SP-1	1277	0.1310	-	7471B	PPB	07 Feb 2024 12:36:27	U	SMPL
P1391-11 - 1	SP-2	1362	0.1475	-	7471B	PPB	07 Feb 2024 12:38:43	U	SMPL
P1391-21 - 1	SP-3	1435	0.1616	-	7471B	PPB	07 Feb 2024 12:40:59	U	SMPL
P1391-31 - 1	SP-4	1384	0.1517	-	7471B	PPB	07 Feb 2024 12:43:17	U	SMPL
P1391-41 - 1	SP-5	1735	0.2198	-	7471B	PPB	07 Feb 2024 12:50:34	U	SMPL
P1391-41DUP - 1	SP-5DUP	1727	0.2182	-	7471B	PPB	07 Feb 2024 12:52:49	U	SMPL
CCV90 - 1	CCV90	24331	4.6015	-	7471B	PPB	07 Feb 2024 12:55:04	U	SMPL
CCB90 - 1	CCB90	797	0.0379	-	7471B	PPB	07 Feb 2024 12:57:20	U	SMPL
P1391-41MS - 1	SP-5MS	20844	3.9253	-	7471B	PPB	07 Feb 2024 12:59:47	U	SMPL
P1391-41MSD - 1	SP-5MSD	19060	3.5793	-	7471B	PPB	07 Feb 2024 13:02:03	U	SMPL
PB158874BL - 1	PBS	351	-0.0486	-	7471B	PPB	07 Feb 2024 13:04:26	U	SMPL
PB158874BS - 1	LCSS	19482	3.6612	-	7471B	PPB	07 Feb 2024 13:06:49	U	SMPL
P1362-01 - 1	WC-1	2264	0.3224	-	7471B	PPB	07 Feb 2024 13:09:04	U	SMPL
P1362-05 - 1	WC-2	10366	1.8935	-	7471B	PPB	07 Feb 2024 13:11:27	U	SMPL
P1362-09 - 1	WC-3	1809	0.2341	-	7471B	PPB	07 Feb 2024 13:13:45	U	SMPL
P1364-01 - 1	COMP-1	4103	0.6790	-	7471B	PPB	07 Feb 2024 13:24:49	U	SMPL
P1364-01DUP - 1	COMP-1DUP	4266	0.7106	-	7471B	PPB	07 Feb 2024 13:27:08	U	SMPL
P1364-01MS - 1	COMP-1MS	24952	4.7219	-	7471B	PPB	07 Feb 2024 13:29:26	U	SMPL
CCV91 - 1	CCV91	27007	5.1204	-	7471B	PPB	07 Feb 2024 13:31:43	U	SMPL
CCB91 - 1	CCB91	687	0.0166	-	7471B	PPB	07 Feb 2024 13:34:09	U	SMPL
P1364-01MSD - 1	COMP-1MSD	24567	4.6472	-	7471B	PPB	07 Feb 2024 13:36:34	U	SMPL
P1364-02 - 1	COMP-2	4692	0.7932	-	7471B	PPB	07 Feb 2024 13:38:50	U	SMPL
P1364-03 - 1	COMP-3	16946	3.1694	-	7471B	PPB	07 Feb 2024 13:41:14	U	SMPL
P1365-01 - 1	B-5(10-12)	17094	3.1981	-	7471B	PPB	07 Feb 2024 13:43:31	U	SMPL
P1366-01 - 1	EB-2-6-8	9161	1.6598	-	7471B	PPB	07 Feb 2024 13:45:52	U	SMPL
P1366-03 - 1	EB-1-0-2	2045	0.2799	-	7471B	PPB	07 Feb 2024 13:48:14	U	SMPL
P1366-05 - 1	EB-1-4-6	2828	0.4317	-	7471B	PPB	07 Feb 2024 13:50:34	U	SMPL
P1367-01 - 1	EB-6-0-2	7055	1.2514	-	7471B	PPB	07 Feb 2024 13:52:52	U	SMPL
P1367-03 - 1	EB-7-0-2	13490	2.4993	-	7471B	PPB	07 Feb 2024 13:55:10	U	SMPL
P1367-05 - 1	EB-6-7-9	610	0.0016	-	7471B	PPB	07 Feb 2024 13:57:28	U	SMPL
CCV92 - 1	CCV92	26944	5.1082	-	7471B	PPB	07 Feb 2024 13:59:47	U	SMPL
CCB92 - 1	CCB92	718	0.0226	-	7471B	PPB	07 Feb 2024 14:02:02	U	SMPL
P1367-07 - 1	EB-7-5-7	18039	3.3814	-	7471B	PPB	07 Feb 2024 14:04:28	U	SMPL
P1367-09 - 1	EB-7-5-7D	29283	5.5617	-	7471B	PPB	07 Feb 2024 14:06:44	U	SMPL
P1367-11 - 1	EB-5-0-2	21888	4.1277	-	7471B	PPB	07 Feb 2024 14:09:07	U	SMPL
P1367-13 - 1	EB-5-6-8	7828	1.4013	-	7471B	PPB	07 Feb 2024 14:11:34	U	SMPL
P1374-01 - 1	BU-03-020524	5650	0.9790	-	7471B	PPB	07 Feb 2024 14:13:57	U	SMPL
P1375-01 - 1	TP-1	38609	7.3702	-	7471B	PPB	07 Feb 2024 14:16:15	U	SMPL
P1391-41LX5 - 1		193	-0.0792	-	7471B	PPB	07 Feb 2024 14:18:34	U	SMPL
P1391-41A - 1		20727	3.9026	-	7471B	PPB	07 Feb 2024 14:21:09	U	SMPL
P1364-01LX5 - 1		577	-0.0048	-	7471B	PPB	07 Feb 2024 14:23:26	U	SMPL
P1364-01A - 1		24497	4.6337	-	7471B	PPB	07 Feb 2024 14:29:08	U	SMPL
CCV93 - 1	CCV93	27076	5.1338	-	7471B	PPB	07 Feb 2024 14:33:55	U	SMPL
CCB93 - 1	CCB93	542	-0.0115	-	7471B	PPB	07 Feb 2024 14:36:15	U	SMPL