



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129423

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/06/2024	10:10
2	CAL2	1	Water	NA	NA	20.2	7.00	02/06/2024	10:11
3	CAL3	1	Water	NA	NA	20.2	10.02	02/06/2024	10:12
4	ICV	1	Water	NA	NA	20.3	7.01	02/06/2024	10:15
5	CCV1	1	Water	NA	NA	20.5	2.02	02/06/2024	10:20
6	P1379-01	1	Solid	20.02	20	21.6	5.15	02/06/2024	10:25
7	P1379-01DUP	1	Solid	20.03	20	21.7	5.16	02/06/2024	10:26
8	CCV2	1	Water	NA	NA	20.3	12.02	02/06/2024	10:30

WORKLIST(Hardcopy Internal Chain)

129423

WorkList Name : PH P1375

WorkList ID : 177666

Department : Wet-Chemistry

Date : 02-06-2024 09:47:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1379-01	RE137-CARBON-20240205	Solid	pH	Cool 4 deg C	TEIR06	D31	02/05/2024	9045D

Date/Time 02/06/24 10:00
 Raw Sample Received by: RM SM
 Raw Sample Relinquished by: RM SM

Date/Time 02/06/24 12:00
 Raw Sample Received by: RM SM
 Raw Sample Relinquished by: RM SM