

Analytical Summary Report

Analysis Method: SM5220 D
 Parameter: COD
 Run Number: LB129427

ANALYST: Iwona
 SUPERVISOR REVIEW BY: Sohil

Reagent/Standard	Lot/Log #
COD calibration std. 10 ppm	WP105791
COD calibration std. 150 ppm	WP105794
COD calibration std. 50 ppm	WP105792
COD calibration std. 0 ppm	WP105790
COD ICV-LCS std, 50ppm	WP105920
COD calibration std. 100 ppm	WP105793
COD ICV-LCS std, 50ppm	WP106454
COD CCV std, 50ppm	WP106453
COD LOD std, 5ppm	WP106455
COD Digestion Vials Low Level 0-150Mg/L	W3067

Temp In (C): <u>148</u>	Date In: <u>02/06/2024</u>	Time In: <u>14:10</u>
Temp Out (C): <u>151</u>	Date Out: <u>02/06/2024</u>	Time Out: <u>16:10</u>

Intercept: 0.0126 Slope: 1.003 Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	-0.013		12/19/2023	14:30
2	CAL2	10	1	Water	11.000	10.955	9.6	12/19/2023	14:30
3	CAL3	50	1	Water	49.000	48.841	-2.3	12/19/2023	14:31
4	CAL4	100	1	Water	100.000	99.688	-0.3	12/19/2023	14:31
5	CAL5	150	1	Water	151.000	150.536	0.4	12/19/2023	14:32



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: Sohil

Run Number: LB129427

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50	NA	NA	1	Water	50.000	49.838	12/19/2023	14:32
2	ICB		NA	NA	1	Water	1.000	0.984	12/19/2023	14:33
3	CCV1	50	NA	NA	1	Water	51.000	50.835	02/07/2024	11:00
4	CCB1		NA	NA	1	Water	0.000	-0.013	02/07/2024	11:00
5	LB129427BL		1	50	1	Solid	1.000	0.984	02/07/2024	11:01
6	LB129427BS	50	1	50	1	Solid	51.000	50.835	02/07/2024	11:01
7	P1187-03		1	50	1	Solid	6.000	5.969	02/07/2024	11:02
8	CCV2	50	NA	NA	1	Water	50.000	49.838	02/07/2024	11:02
9	CCB2		NA	NA	1	Water	1.000	0.984	02/07/2024	11:03