



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129430

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.7	02/07/2024	13:35
2	P1373-01		1	12	100.00	>212.0	>212.0	02/07/2024	14:00
3	P1373-01DUP		1	12	100.00	>212.0	>212.0	02/07/2024	14:30
4	P1373-02		1	14	100.00	>212.0	>212.0	02/07/2024	15:00
5	P1373-03		1	13	100.00	>212.0	>212.0	02/07/2024	15:30
6	P1373-04		1	12	100.00	>212.0	>212.0	02/07/2024	16:00
7	P1373-05		1	12	100.00	>212.0	>212.0	02/07/2024	16:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16129430-

WorkList Name : fp0206

WorkList ID : 177692

Department : Wet-Chemistry

Date : 02-06-2024 09:49:00

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1373-01	1289	Water	Flash Point	Cool 4 deg C	PSEG03	D31	02/05/2024	1010B
P1373-02	1290	Water	Flash Point	Cool 4 deg C	PSEG03	D31	02/05/2024	1010B
P1373-03	1705	Water	Flash Point	Cool 4 deg C	PSEG03	D31	02/05/2024	1010B
P1373-04	1154	Water	Flash Point	Cool 4 deg C	PSEG03	D31	02/05/2024	1010B
P1373-05	HT-020524	Water	Flash Point	Cool 4 deg C	PSEG03	D31	02/05/2024	1010B

Date/Time 02/07/2024 13:30
Raw Sample Received by: RM (Case)
Raw Sample Relinquished by: TD (Case)

Date/Time 02/07/2024 16:46
Raw Sample Received by: TD (Case)
Raw Sample Relinquished by: RM (Case)