



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB129431

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/07/2024	15:10
2	CAL2	1	Water	NA	NA	20.2	7.00	02/07/2024	15:11
3	CAL3	1	Water	NA	NA	20.3	10.02	02/07/2024	15:15
4	ICV	1	Water	NA	NA	20.4	7.02	02/07/2024	15:20
5	CCV1	1	Water	NA	NA	20.3	2.01	02/07/2024	15:22
6	P1388-02	1	Water	NA	NA	20.1	7.45	02/07/2024	15:33
7	P1388-02DUP	1	Water	NA	NA	20.2	7.46	02/07/2024	15:34
8	P1388-03	1	Water	NA	NA	20.5	7.18	02/07/2024	15:38
9	P1388-04	1	Water	NA	NA	20.2	7.09	02/07/2024	15:40
10	P1388-05	1	Water	NA	NA	20.3	7.04	02/07/2024	15:45
11	CCV2	1	Water	NA	NA	20.3	12.02	02/07/2024	15:50