


LB129432

SFAM01.1

INSTRUMENT ID: CV1

 μ Abs.:

49907



Linear

A= 0.0000e+000

B= 2.0010e-004 slope

C= -6.9229e-002 $\frac{1}{\text{slope}}$ intercept

Rho= 0.9995809

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/p
0.00	0.000	-0.009	-0.009	299	0.000	299					=
0.05	0.050					442	0				-9
0.20	0.200	0.183	-0.017	1258	0.0 %	1258					-4
2.50	2.500	2.399	-0.101	12336	0.0 %	12336					5
5.00	5.000	5.226	0.226	26460	0.0 %	26460					0
7.50	7.500	7.485	-0.015	37750	0.0 %	37750					-1
10.0	10.000	9.917	-0.083	49907	0.0 %	49907					

LB129432
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	299	-		0 SFAM01.1	PPB	2/7/2024 15:26	Std
	0.2 S01	1258	-		0.2 SFAM01.1	PPB	2/7/2024 15:28	Std
	2.5 S02	12336	-		2.5 SFAM01.1	PPB	2/7/2024 15:31	Std
	5 S03	26460	-		5 SFAM01.1	PPB	2/7/2024 15:33	Std
	7.5 S04	37750	-		7.5 SFAM01.1	PPB	2/7/2024 15:35	Std
	10 S05	49907	-		10 SFAM01.1	PPB	2/7/2024 15:37	Std
ICV099	ICV099	20134	3.9597 -		SFAM01.1	PPB	2/7/2024 15:40	SMPL
ICB099	ICB099	129	-0.0434 -		SFAM01.1	PPB	2/7/2024 15:43	SMPL
CCV094	CCV094	26893	5.3122 -		SFAM01.1	PPB	2/7/2024 15:45	SMPL
CCB094	CCB094	86	-0.052 -		SFAM01.1	PPB	2/7/2024 15:47	SMPL
PB158886BL	PBW886	313	-0.0066 -		SFAM01.1	PPB	2/7/2024 15:49	SMPL
P1216-01	YBRR9	731	0.077 -		SFAM01.1	PPB	2/7/2024 15:52	SMPL
P1216-02	YBRR9D	814	0.0937 -		SFAM01.1	PPB	2/7/2024 15:54	SMPL
P1216-03	YBRR9S	5613	1.054 -		SFAM01.1	PPB	2/7/2024 15:59	SMPL
P1216-04	YBR50	920	0.1149 -		SFAM01.1	PPB	2/7/2024 16:01	SMPL
P1216-05	YBR51	2337	0.3984 -		SFAM01.1	PPB	2/7/2024 16:03	SMPL
P1216-06	YBR52	1467	0.2243 -		SFAM01.1	PPB	2/7/2024 16:06	SMPL
P1216-07	YBR53	1506	0.2321 -		SFAM01.1	PPB	2/7/2024 16:08	SMPL
P1216-08	YBR54	1219	0.1747 -		SFAM01.1	PPB	2/7/2024 16:10	SMPL
P1216-09	YBR55	1003	0.1315 -		SFAM01.1	PPB	2/7/2024 16:12	SMPL
P1253-01	MBH6Z9	248	-0.0196 -		SFAM01.1	PPB	2/7/2024 16:15	SMPL
P1253-02	MBH703	277	-0.0138 -		SFAM01.1	PPB	2/7/2024 16:17	SMPL
P1253-03	MBH703D	306	-0.008 -		SFAM01.1	PPB	2/7/2024 16:19	SMPL
P1253-04	MBH703S	5297	0.9907 -		SFAM01.1	PPB	2/7/2024 16:24	SMPL
P1253-05	MBH704	275	-0.0142 -		SFAM01.1	PPB	2/7/2024 16:26	SMPL
P1266-05	MJPTA2	268	-0.0156 -		SFAM01.1	PPB	2/7/2024 16:29	SMPL
P1266-06	MJPTA3	288	-0.0116 -		SFAM01.1	PPB	2/7/2024 16:31	SMPL
P1266-07	MJPTA3D	279	-0.0134 -		SFAM01.1	PPB	2/7/2024 16:33	SMPL
P1266-08	MJPTA3S	5330	0.9973 -		SFAM01.1	PPB	2/7/2024 16:35	SMPL
CCV095	CCV095	26242	5.1819 -		SFAM01.1	PPB	2/7/2024 16:38	SMPL
CCB095	CCB095	89	-0.0514 -		SFAM01.1	PPB	2/7/2024 16:40	SMPL