



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB129433

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	02/07/2024	15:45
2	CAL2	1	Water	NA	NA	20.4	6.99	02/07/2024	15:46
3	CAL3	1	Water	NA	NA	20.3	10.02	02/07/2024	15:48
4	ICV	1	Water	NA	NA	20.4	6.98	02/07/2024	15:50
5	CCV1	1	Water	NA	NA	20.5	2.01	02/07/2024	16:00
6	P1339-02	1	Solid	20.03	20	22.9	10.22	02/07/2024	16:20
7	P1339-02DUP	1	Solid	20.04	20	22.8	10.23	02/07/2024	16:21
8	P1411-03	1	Solid	20.02	20	22.3	6.87	02/07/2024	16:25
9	P1412-02	1	Solid	20.01	20	23.0	7.52	02/07/2024	16:30
10	CCV2	1	Water	NA	NA	20.3	12.02	02/07/2024	16:33

WORKLIST(Hardcopy Internal Chain)

VP 129433

WorkList Name : corrosivity p1411 WorkList ID : 177716 Department : Wet-Chemistry Date : 02-07-2024 15:20:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1339-02	WC-COMP	Solid	Corrosivity	Cool 4 deg C	EART12	D11	02/01/2024	9045D
P1411-03	72-11929	Solid	Corrosivity	Cool 4 deg C	PSEG03	E31	02/07/2024	9045D
P1412-02	TP2	Solid	Corrosivity	Cool 4 deg C	PSEG03	E31	02/07/2024	9045D

Date/Time 02/07/24 15:20

Raw Sample Received by: SA (CSC)

Raw Sample Relinquished by: IT (CSC)

Date/Time 02/07/24 17:30

Raw Sample Received by: IT (CSC)

Raw Sample Relinquished by: SA (CSC)