



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129434

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	02/07/2024	15:45
2	CAL2	1	Water	NA	NA	20.4	6.99	02/07/2024	15:46
3	CAL3	1	Water	NA	NA	20.3	10.02	02/07/2024	15:48
4	ICV	1	Water	NA	NA	20.4	6.98	02/07/2024	15:50
5	CCV1	1	Water	NA	NA	20.5	2.01	02/07/2024	16:00
6	P1304-01	1	Solid	20.02	20	20.2	7.44	02/07/2024	16:11
7	P1339-01	1	Solid	20.03	20	22.9	10.22	02/07/2024	16:20
8	P1339-01DUP	1	Solid	20.04	20	22.8	10.23	02/07/2024	16:21
9	CCV2	1	Water	NA	NA	20.4	12.02	02/07/2024	16:25

WORKLIST(Hardcopy Internal Chain)

UB 129434

WorkList Name : ph s p1304 WorkList ID : 177715 Department : Wet-Chemistry Date : 02-07-2024 14:56:45

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1304-01	BRIGHTON-WEST-8TH-STREE	Solid	pH	Cool 4 deg C	TULL02	D11	01/30/2024	9045D
P1339-01	WC-COMP	Solid	pH	Cool 4 deg C	EART12	D11	02/01/2024	9045D

Date/Time 0210724 15:10
 Raw Sample Received by: SA COCI
 Raw Sample Relinquished by: Jm

Date/Time 0210724 17:30
 Raw Sample Received by: Jm
 Raw Sample Relinquished by: SA COCI