


INSTRUMENT IDICW

INSTRUMENT IDICW

41394



A= 0.0000e+000

B= 2.3749e-004

C= -6.6175e-002

Rho= 0.9980179

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.004	-0.004	262	0.000	262				
0.2	0.200	0.205	0.005	1140	0.0 %	1140	0			
2.5	2.500	2.586	0.086	11166	0.0 %	11166				
5.0	5.000	4.721	-0.279	20158	0.0 %	20158				
7.5	7.500	7.928	0.428	33660	0.0 %	33660				
10.0	10.000	9.765	-0.235	41394	0.0 %	41394				

LB129440 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Feb 2024 10:28:52

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	262	-	0.0000	7470A	PPB	08 Feb 2024 10:39:12	S	Std
0.2 - 1	50.2	1140	-	0.2000	7470A	PPB	08 Feb 2024 10:41:29	S	Std
2.5 - 1	52.5	11166	-	2.5000	7470A	PPB	08 Feb 2024 10:43:47	S	Std
5.0 - 1	55	20158	-	5.0000	7470A	PPB	08 Feb 2024 10:46:04	S	Std
7.5 - 1	57.5	33660	-	7.5000	7470A	PPB	08 Feb 2024 10:48:23	S	Std
10.0 - 1	57.5	41394	-	10.0000	7470A	PPB	08 Feb 2024 10:50:41	S	Std
ICV100 - 1	ICV100	16369	3.8214	-	7470A	PPB	08 Feb 2024 10:53:26	U	SMPL
ICB100 - 1	ICB100	140	-0.0329	-	7470A	PPB	08 Feb 2024 10:55:42	U	SMPL
CCV96 - 1	CCV96	20351	4.7671	-	7470A	PPB	08 Feb 2024 10:57:59	U	SMPL
CCB96 - 1	CCB96	120	-0.0377	-	7470A	PPB	08 Feb 2024 11:00:14	U	SMPL
CRA - 1	CRA	1065	0.1868	-	7470A	PPB	08 Feb 2024 11:02:33	U	SMPL
HighStd - 1	HighStd	40822	9.6288	-	7470A	PPB	08 Feb 2024 11:04:47	U	SMPL
ChkStd - 1	ChkStd	30846	7.2595	-	7470A	PPB	08 Feb 2024 11:10:13	U	SMPL
PB158880BL - 1	PBW	121	-0.0374	-	7470A	PPB	08 Feb 2024 11:19:00	U	SMPL
PB158880BS - 1	LCSW	17340	4.0520	-	7470A	PPB	08 Feb 2024 11:21:16	U	SMPL
P1375-02 - 1	TP-1	-138	-0.0989	-	7470A	PPB	08 Feb 2024 11:23:34	U	SMPL
P1378-02 - 1	TP-4	-4	-0.0671	-	7470A	PPB	08 Feb 2024 11:25:53	U	SMPL
P1378-04 - 1	TP-5	8	-0.0643	-	7470A	PPB	08 Feb 2024 11:28:10	U	SMPL
P1379-02 - 1	RE137-CARBON-20240205	78	-0.0477	-	7470A	PPB	08 Feb 2024 11:30:27	U	SMPL
P1389-04 - 1	WC-1	110	-0.0401	-	7470A	PPB	08 Feb 2024 11:32:42	U	SMPL
CCV97 - 1	CCV97	20633	4.8340	-	7470A	PPB	08 Feb 2024 11:34:57	U	SMPL
CCB97 - 1	CCB97	137	-0.0336	-	7470A	PPB	08 Feb 2024 11:37:13	U	SMPL
P1389-08 - 1	WC-2	94	-0.0439	-	7470A	PPB	08 Feb 2024 11:39:31	U	SMPL
P1389-12 - 1	WC-3	105	-0.0412	-	7470A	PPB	08 Feb 2024 11:41:47	U	SMPL
P1389-16 - 1	WC-4	149	-0.0308	-	7470A	PPB	08 Feb 2024 11:44:02	U	SMPL
P1391-10 - 1	SP-1	168	-0.0263	-	7470A	PPB	08 Feb 2024 11:46:18	U	SMPL
P1391-20 - 1	SP-2	101	-0.0422	-	7470A	PPB	08 Feb 2024 11:48:34	U	SMPL
P1391-30 - 1	SP-3	57	-0.0526	-	7470A	PPB	08 Feb 2024 11:50:50	U	SMPL
P1391-40 - 1	SP-4	77	-0.0479	-	7470A	PPB	08 Feb 2024 11:53:07	U	SMPL
P1391-50 - 1	SP-5	105	-0.0412	-	7470A	PPB	08 Feb 2024 11:56:59	U	SMPL
P1391-50DUP - 1	SP-5DUP	122	-0.0372	-	7470A	PPB	08 Feb 2024 11:59:17	U	SMPL
P1391-50MS - 1	SP-5MS	17080	3.9902	-	7470A	PPB	08 Feb 2024 12:03:50	U	SMPL
CCV98 - 1	CCV98	20368	4.7711	-	7470A	PPB	08 Feb 2024 12:06:06	U	SMPL
CCB98 - 1	CCB98	64	-0.0510	-	7470A	PPB	08 Feb 2024 12:08:24	U	SMPL
P1391-50MSD - 1	SP-5MSD	17173	4.0123	-	7470A	PPB	08 Feb 2024 12:10:42	U	SMPL
PB158849TB - 1	PB158849TB	-132	-0.0975	-	7470A	PPB	08 Feb 2024 12:12:57	U	SMPL
P1391-50LX5 - 1		-31	-0.0735	-	7470A	PPB	08 Feb 2024 12:15:15	U	SMPL
P1391-50A - 1		16591	3.8741	-	7470A	PPB	08 Feb 2024 12:17:31	U	SMPL
CCV99 - 1	CCV99	20344	4.7654	-	7470A	PPB	08 Feb 2024 12:19:46	U	SMPL
CCB99 - 1	CCB99	52	-0.0538	-	7470A	PPB	08 Feb 2024 12:22:05	U	SMPL