



Analytical Summary Report

Analysis Method: SM2120 B
Parameter: Color
Run Number: LB129449

Reviewed By: Iwona
Supervisor Review By: Sohil
pH Meter ID: WC pH METER-1

Reagent/Standard	Lot/Log #
Calibration std, 0 color unit	WP106021
Calibration std, 5 color unit	WP106022
Calibration std, 10 color unit	WP106023
Calibration std, 15 color unit	WP106024
Calibration std, 20 color unit	WP106025
Calibration std, 25 color unit	WP106026
Calibration std, 30 color unit	WP106027
Calibration std, 40 color unit	WP106028
Calibration std, 50 color unit	WP106029
Calibration std, 60 color unit	WP106030
Calibration std, 70 color unit	WP106031
Calibration std, 100 color unit	WP106032

Seq	LabID	ClientID	DL	pH Reading	pH Adjustment	Result (cu)	Anal Date	Anal Time
1	LB129449BL	LB129449BL	1	5.03	N/A	0.00	02/08/2024	14:40
2	P1404-01	MONTHLY-OUTFALL-002	10	8.09	N/A	60.00	02/08/2024	14:45
3	P1404-01DUP	MONTHLY-OUTFALL-002D	10	8.07	N/A	60.00	02/08/2024	14:50

WORKLIST(Hardcopy Internal Chain)

6129449

WorkList Name : COLOR-020824

WorkList ID : 177746

Department : Wet-Chemistry

Date : 02-08-2024 14:26:54

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1404-01	MONTHLY-OUTFALL-002	Water	Color	Cool 4 deg C	INFR01	E31	02/07/2024	SM2120 B

Date/Time 02/08/24 14:30
Raw Sample Received by: 12(50)
Raw Sample Relinquished by: 12(50)

Date/Time 02/08/24 15:00
Raw Sample Received by: 12(50)
Raw Sample Relinquished by: 12(50)