



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129455

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/08/2024	16:10
2	CAL2	1	Water	NA	NA	20.3	7.00	02/08/2024	16:11
3	CAL3	1	Water	NA	NA	20.3	10.02	02/08/2024	16:12
4	ICV	1	Water	NA	NA	20.4	7.02	02/08/2024	16:15
5	CCV1	1	Water	NA	NA	20.3	2.02	02/08/2024	16:45
6	P1422-01	1	Water	NA	NA	20.4	4.73	02/08/2024	16:50
7	P1427-01	1	Water	NA	NA	20.1	5.46	02/08/2024	17:00
8	P1427-02	1	Water	NA	NA	20.2	5.59	02/08/2024	17:05
9	P1427-02DUP	1	Water	NA	NA	20.2	5.60	02/08/2024	17:06
10	CCV2	1	Water	NA	NA	20.3	12.02	02/08/2024	17:10

15129455

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p1422

WorkList ID : 177756

Department : Wet-Chemistry

Date : 02-08-2024 15:57:35

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1422-01	DRUM-1A	Water	pH	Cool 4 deg C	PSEG03	E21	02/08/2024	9040C
P1427-01	50239	Water	pH	Cool 4 deg C	PSEG03	E21	02/08/2024	9040C
P1427-02	50240	Water	pH	Cool 4 deg C	PSEG03	E21	02/08/2024	9040C

Date/Time 02/08/24 16:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 02/08/24 18:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]