

LB129470


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51926



A= 0.0000e+000

B= 1.8757e-004

C= -1.1976e-001

Rho= 0.9977041

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/0D
0.0	0.000	-0.100	-0.100	105	0.000	105					2
0.2	0.200	0.205	0.005	1731	0.0 %	1731					2
2.5	2.500	2.546	0.046	14213	0.0 %	14213					2
5.0	5.000	4.963	-0.037	27100	0.0 %	27100					-1
7.5	7.500	7.966	0.466	43106	0.0 %	43106					6
10.0	10.000	9.620	-0.380	51926	0.0 %	51926					-4

LB129470 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 09 Feb 2024 13:28:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	105	-	0.0000	7471B	PPB	09 Feb 2024 13:35:14	S	Std
0.2 - 1	58-2	1731	-	0.2000	7471B	PPB	09 Feb 2024 13:37:31	S	Std
2.5 - 1	52.2	14213	-	2.5000	7471B	PPB	09 Feb 2024 13:39:48	S	Std
5.0 - 1	35	27100	-	5.0000	7471B	PPB	09 Feb 2024 13:46:03	S	Std
7.5 - 1	35	43106	-	7.5000	7471B	PPB	09 Feb 2024 13:48:19	S	Std
10.0 - 1	35	51926	-	10.0000	7471B	PPB	09 Feb 2024 13:50:51	S	Std
ICV03 - 1	ICV03	22306	4.0642	-	7471B	PPB	09 Feb 2024 13:56:37	U	SMPL
ICB03 - 1	ICB03	250	-0.0729	-	7471B	PPB	09 Feb 2024 13:58:55	U	SMPL
CCV07 - 1	CCV07	27166	4.9758	-	7471B	PPB	09 Feb 2024 14:01:16	U	SMPL
CCB07 - 1	CCB07	252	-0.0725	-	7471B	PPB	09 Feb 2024 14:03:32	U	SMPL
CRA - 1	CRA	1698	0.1987	-	7471B	PPB	09 Feb 2024 14:05:57	U	SMPL
HighStd - 1	HighStd	51407	9.5226	-	7471B	PPB	09 Feb 2024 14:08:13	U	SMPL
ChkStd - 1	ChkStd	39045	7.2039	-	7471B	PPB	09 Feb 2024 14:10:28	U	SMPL
PB158946BL - 1	PBS	183	-0.0854	-	7471B	PPB	09 Feb 2024 14:13:01	U	SMPL
PB158946BS - 1	LCSS	21971	4.0013	-	7471B	PPB	09 Feb 2024 14:15:32	U	SMPL
P1387-01 - 1	HD-01-020724	1746	0.2077	-	7471B	PPB	09 Feb 2024 14:22:34	U	SMPL
P1387-01DUP - 1	HD-01-020724DUP	1595	0.1794	-	7471B	PPB	09 Feb 2024 14:24:51	U	SMPL
P1387-01MS - 1	HD-01-020724MS	26174	4.7897	-	7471B	PPB	09 Feb 2024 14:27:09	U	SMPL
P1387-01MSD - 1	HD-01-020724MSD	25534	4.6696	-	7471B	PPB	09 Feb 2024 14:29:26	U	SMPL
P1389-01 - 1	WC-1	4900	0.7993	-	7471B	PPB	09 Feb 2024 14:31:50	U	SMPL
CCV08 - 1	CCV08	25867	4.7321	-	7471B	PPB	09 Feb 2024 14:34:16	U	SMPL
CCB08 - 1	CCB08	203	-0.0817	-	7471B	PPB	09 Feb 2024 14:36:34	U	SMPL
P1389-05 - 1	WC-2	2369	0.3246	-	7471B	PPB	09 Feb 2024 14:38:59	U	SMPL
P1389-09 - 1	WC-3	5156	0.8473	-	7471B	PPB	09 Feb 2024 14:41:14	U	SMPL
P1389-13 - 1	WC-4	3992	0.6290	-	7471B	PPB	09 Feb 2024 14:43:30	U	SMPL
P1411-01 - 1	72-11929	6399	1.0805	-	7471B	PPB	09 Feb 2024 14:45:48	U	SMPL
P1413-01 - 1	VNJ-209	11497	2.0367	-	7471B	PPB	09 Feb 2024 14:48:07	U	SMPL
P1413-03 - 1	ETGI-347	4704	0.7626	-	7471B	PPB	09 Feb 2024 14:50:26	U	SMPL
P1413-05 - 1	ETGI-358	4587	0.7406	-	7471B	PPB	09 Feb 2024 14:52:46	U	SMPL
P1428-01 - 1	EO-01-020824	3951	0.6213	-	7471B	PPB	09 Feb 2024 14:55:05	U	SMPL
P1435-01 - 1	FRENCH-020724	5135	0.8434	-	7471B	PPB	09 Feb 2024 14:57:24	U	SMPL
CCV09 - 1	CCV09	27043	4.9527	-	7471B	PPB	09 Feb 2024 15:07:13	U	SMPL
CCB09 - 1	CCB09	374	-0.0496	-	7471B	PPB	09 Feb 2024 15:09:28	U	SMPL
P1387-01LX5 - 1		311	-0.0614	-	7471B	PPB	09 Feb 2024 15:11:54	U	SMPL
P1387-01A - 1		22514	4.1032	-	7471B	PPB	09 Feb 2024 15:14:09	U	SMPL
CCV10 - 1	CCV10	28878	5.2969	-	7471B	PPB	09 Feb 2024 15:16:25	U	SMPL
CCB10 - 1	CCB10	457	-0.0340	-	7471B	PPB	09 Feb 2024 15:18:50	U	SMPL