



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129472

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/09/2024	12:15
2	CAL2	1	Water	NA	NA	20.2	7.00	02/09/2024	12:16
3	CAL3	1	Water	NA	NA	20.3	10.02	02/09/2024	12:19
4	ICV	1	Water	NA	NA	20.4	7.02	02/09/2024	12:20
5	CCV1	1	Water	NA	NA	20.3	2.01	02/09/2024	12:22
6	P1441-01	1	Water	NA	NA	20.1	4.60	02/09/2024	12:25
7	P1441-01DUP	1	Water	NA	NA	20.2	4.61	02/09/2024	12:26
8	P1441-02	1	Water	NA	NA	20.2	4.93	02/09/2024	12:30
9	P1441-03	1	Water	NA	NA	20.2	4.96	02/09/2024	12:33
10	CCV2	1	Water	NA	NA	20.3	12.02	02/09/2024	12:36

120912

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p1441 WorkList ID : 177783 Department : Wet-Chemistry Date : 02-09-2024 12:04:07

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1441-01	CELL-A-POOL-WATER	Water	pH	Cool 4 deg C	STER13	E21	02/08/2024	9040C
P1441-02	CELL-B-POOL-WATER	Water	pH	Cool 4 deg C	STER13	E21	02/08/2024	9040C
P1441-03	CELL-C-POOL-WATER	Water	pH	Cool 4 deg C	STER13	E21	02/08/2024	9040C

Date/Time 02/09/24 12:10
Raw Sample Received by: SP (WC)
Raw Sample Relinquished by: CP SM

Date/Time 02/09/24 12:00
Raw Sample Received by: CP SM
Raw Sample Relinquished by: SP (WC)