



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB129480

Slope : 97.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/09/2024	16:25
2	CAL2	1	Water	NA	NA	20.3	7.00	02/09/2024	16:26
3	CAL3	1	Water	NA	NA	20.3	10.02	02/09/2024	16:30
4	ICV	1	Water	NA	NA	20.4	7.01	02/09/2024	16:35
5	CCV1	1	Water	NA	NA	20.4	2.01	02/09/2024	16:44
6	P1446-02	1	Solid	20.02	20	23.8	11.70	02/09/2024	16:50
7	P1446-02DUP	1	Solid	20.03	20	23.9	11.71	02/09/2024	16:51
8	P1449-02	1	Solid	20.02	20	23.9	8.20	02/09/2024	17:00
9	P1453-04	1	Solid	20.03	20	23.8	8.24	02/09/2024	17:10
10	P1453-08	1	Solid	20.04	20	24.2	8.17	02/09/2024	17:15
11	CCV2	1	Water	NA	NA	20.4	12.02	02/09/2024	17:20

UP 29480

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity p1446

WorkList ID : 177788

Department : Wet-Chemistry

Date : 02-09-2024 15:53:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1446-02	RT3475	Solid	Corrosivity	Cool 4 deg C	PSEG03	E21	02/09/2024	9045D
P1449-02	ETGI-362	Solid	Corrosivity	Cool 4 deg C	PSEG03	E21	02/09/2024	9045D
P1453-04	TP-3	Solid	Corrosivity	Cool 4 deg C	PSEG03	E21	02/09/2024	9045D
P1453-08	TP-2	Solid	Corrosivity	Cool 4 deg C	PSEG03	E21	02/09/2024	9045D

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

02/09/24 16:00

JP CWC

JP SPM

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

02/09/24

JP SPM

JP CWC

18:00