

L3129486


7470A

INSTRUMENT ID: CY2

Linear

μ Abs.:

29930



A= 0.0000e+000

B= 3.3490e-004

C= -1.4613e-001

Rho= 0.9996994

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/1))
0.0	0.000	-0.107	-0.107	117	0.000	117					
0.2	0.200	0.204	0.004	1046	0.0 %	1046	0				2
2.5	2.500	2.596	0.096	8187	0.0 %	8187					4
5.0	5.000	5.014	0.014	15409	0.0 %	15409					0
7.5	7.500	7.615	0.115	23175	0.0 %	23175					2
10.0	10.000	9.877	-0.123	29930	0.0 %	29930					2

LB129486 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Feb 2024 10:31:28

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	117	-	0.0000	7470A	PPB	12 Feb 2024 10:45:51	S	Std
0.2 - 1	50.2	1046	-	0.2000	7470A	PPB	12 Feb 2024 10:48:08	S	Std
2.5 - 1	52.5	8187	-	2.5000	7470A	PPB	12 Feb 2024 10:50:25	S	Std
5.0 - 1	55	15409	-	5.0000	7470A	PPB	12 Feb 2024 10:52:41	S	Std
7.5 - 1	57.5	23175	-	7.5000	7470A	PPB	12 Feb 2024 10:54:59	S	Std
10.0 - 1	510	29930	-	10.0000	7470A	PPB	12 Feb 2024 10:59:55	S	Std
ICV04 - 1	ICV04	11990	3.8693	-	7470A	PPB	12 Feb 2024 11:02:52	U	SMPL
ICB04 - 1	ICB04	349	-0.0293	-	7470A	PPB	12 Feb 2024 11:05:07	U	SMPL
CCV11 - 1	CCV11	15604	5.0797	-	7470A	PPB	12 Feb 2024 11:12:09	U	SMPL
CCB11 - 1	CCB11	597	0.0538	-	7470A	PPB	12 Feb 2024 11:14:24	U	SMPL
CRA - 1	CRA	991	0.1858	-	7470A	PPB	12 Feb 2024 11:19:13	U	SMPL
HighStd - 1	HighStd	27216	8.9686	-	7470A	PPB	12 Feb 2024 11:24:08	U	SMPL
ChkStd - 1	ChkStd	20900	6.8533	-	7470A	PPB	12 Feb 2024 11:26:25	U	SMPL
PB158931BL - 1	PBW	357	-0.0266	-	7470A	PPB	12 Feb 2024 11:28:44	U	SMPL
PB158931BS - 1	LCSW	12054	3.8908	-	7470A	PPB	12 Feb 2024 11:34:00	U	SMPL
P1339-02 - 1	WC-COMP	558	0.0407	-	7470A	PPB	12 Feb 2024 11:36:15	U	SMPL
P1411-03 - 1	72-11929	532	0.0320	-	7470A	PPB	12 Feb 2024 11:42:09	U	SMPL
P1412-02 - 1	TP2	757	0.1074	-	7470A	PPB	12 Feb 2024 11:44:25	U	SMPL
P1412-02DUP - 1	TP2DUP	213	-0.0748	-	7470A	PPB	12 Feb 2024 11:46:42	U	SMPL
P1412-02MS - 1	TP2MS	12982	4.2016	-	7470A	PPB	12 Feb 2024 11:53:07	U	SMPL
CCV12 - 1	CCV12	14506	4.7120	-	7470A	PPB	12 Feb 2024 11:55:24	U	SMPL
CCB12 - 1	CCB12	528	0.0307	-	7470A	PPB	12 Feb 2024 11:57:44	U	SMPL
P1412-02MSD - 1	TP2MSD	12375	3.9983	-	7470A	PPB	12 Feb 2024 12:00:02	U	SMPL
PB158961BL - 1	PBW	154	-0.0946	-	7470A	PPB	12 Feb 2024 12:07:22	U	SMPL
PB158961BS - 1	LCSW	12693	4.1048	-	7470A	PPB	12 Feb 2024 12:12:15	U	SMPL
P1422-01 - 1	DRUM-1A	423	-0.0045	-	7470A	PPB	12 Feb 2024 12:14:30	U	SMPL
P1422-01DUP - 1	DRUM-1ADUP	372	-0.0216	-	7470A	PPB	12 Feb 2024 12:21:38	U	SMPL
P1422-01MS - 1	DRUM-1AMS	10605	3.4055	-	7470A	PPB	12 Feb 2024 12:23:54	U	SMPL
P1422-01MSD - 1	DRUM-1AMSD	12003	3.8737	-	7470A	PPB	12 Feb 2024 12:26:10	U	SMPL
PB158969BL - 1	PBW	476	0.0133	-	7470A	PPB	12 Feb 2024 12:31:13	U	SMPL
PB158969BS - 1	LCSW	11695	3.7705	-	7470A	PPB	12 Feb 2024 12:35:47	U	SMPL
P1368-03 - 1	MR-306P-GW-20240201-0	13091	4.2381	-	7470A	PPB	12 Feb 2024 12:38:05	U	SMPL
CCV13 - 1	CCV13	14172	4.6001	-	7470A	PPB	12 Feb 2024 12:40:23	U	SMPL
CCB13 - 1	CCB13	461	0.0083	-	7470A	PPB	12 Feb 2024 12:42:42	U	SMPL
P1373-05 - 1	HT-020524	2651	0.7417	-	7470A	PPB	12 Feb 2024 12:45:01	U	SMPL
P1373-05DUP - 1	HT-020524DUP	2605	0.7263	-	7470A	PPB	12 Feb 2024 12:47:16	U	SMPL
P1373-05MS - 1	HT-020524MS	13357	4.3272	-	7470A	PPB	12 Feb 2024 12:57:03	U	SMPL
P1373-05MSD - 1	HT-020524MSD	15129	4.9206	-	7470A	PPB	12 Feb 2024 12:59:19	U	SMPL
P1407-02 - 1	COMP	737	0.1007	-	7470A	PPB	12 Feb 2024 13:01:37	U	SMPL
P1414-09 - 1	FB020724	227	-0.0701	-	7470A	PPB	12 Feb 2024 13:06:39	U	SMPL
PB158884TB - 1	PB158884TB	225	-0.0708	-	7470A	PPB	12 Feb 2024 13:11:33	U	SMPL
PB158914TB - 1	PB158914TB	288	-0.0497	-	7470A	PPB	12 Feb 2024 13:13:49	U	SMPL
CCV14 - 1	CCV14	14733	4.7880	-	7470A	PPB	12 Feb 2024 13:16:07	U	SMPL
CCB14 - 1	CCB14	343	-0.0313	-	7470A	PPB	12 Feb 2024 13:21:00	U	SMPL
P1412-02LX5 - 1		235	-0.0674	-	7470A	PPB	12 Feb 2024 13:23:15	U	SMPL
P1412-02A - 1		14486	4.7053	-	7470A	PPB	12 Feb 2024 13:25:31	U	SMPL
P1422-01LX5 - 1		-63	-0.1672	-	7470A	PPB	12 Feb 2024 13:27:46	U	SMPL
P1422-01A - 1		12410	4.0100	-	7470A	PPB	12 Feb 2024 13:30:04	U	SMPL
P1373-05LX5 - 1		616	0.0602	-	7470A	PPB	12 Feb 2024 13:32:20	U	SMPL
P1373-05A - 1		15653	5.0961	-	7470A	PPB	12 Feb 2024 13:34:38	U	SMPL
CCV15 - 1	CCV15	14489	4.7063	-	7470A	PPB	12 Feb 2024 13:36:54	U	SMPL
CCB15 - 1	CCB15	244	-0.0644	-	7470A	PPB	12 Feb 2024 13:39:14	U	SMPL