



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129489

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	02/12/2024	12:20
2	P1388-01		1	12	100.00	>212.0	>212.0	02/12/2024	12:50
3	P1388-01DUP		1	14	100.00	>212.0	>212.0	02/12/2024	13:20
4	P1422-01		1	13	100.00	>212.0	>212.0	02/12/2024	13:50
5	P1424-01		1	14	100.00	>212.0	>212.0	02/12/2024	14:20
6	P1424-02		1	13	100.00	>212.0	>212.0	02/12/2024	14:50
7	P1424-03		1	12	100.00	>212.0	>212.0	02/12/2024	15:20
8	P1427-01		1	14	100.00	>212.0	>212.0	02/12/2024	15:50
9	P1427-02		1	13	100.00	>212.0	>212.0	02/12/2024	16:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-129489

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 02/12/24 flashpoint w WorkList ID : 177831 Department : Wet-Chemistry Date : 02-12-2024 09:05:39

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1388-01	60135-60136-60137-60138-COI	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/07/2024	1010B
P1422-01	DRUM-1A	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/08/2024	1010B
P1424-01	1281	Water	Flash Point	Cool 4 deg C	PSEG03	E22	02/08/2024	1010B
P1424-02	1285	Water	Flash Point	Cool 4 deg C	PSEG03	E22	02/08/2024	1010B
P1424-03	1288	Water	Flash Point	Cool 4 deg C	PSEG03	E22	02/08/2024	1010B
P1427-01	50239	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/08/2024	1010B
P1427-02	50240	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/08/2024	1010B

Date/Time 02/12/2024 12:10
Raw Sample Received by: N.P. W.C.
Raw Sample Relinquished by: [Signature]

Date/Time 02/12/2024 16:55
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: N.P. W.C.