



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129494

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/12/2024	16:05
2	CAL2	1	Water	NA	NA	20.3	7.00	02/12/2024	16:06
3	CAL3	1	Water	NA	NA	20.3	10.02	02/12/2024	16:09
4	ICV	1	Water	NA	NA	20.4	7.01	02/12/2024	16:10
5	CCV1	1	Water	NA	NA	20.3	2.01	02/12/2024	16:11
6	P1458-01	1	Water	NA	NA	20.7	6.96	02/12/2024	16:15
7	CCV2	1	Water	NA	NA	20.3	12.03	02/12/2024	16:19

WORKLIST(Hardcopy Internal Chain)

UP 129944

WorkList Name : ph p1458 water

WorkList ID : 177836

Department : Wet-Chemistry

Date : 02-12-2024 15:59:40

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1458-01	UP	Water	pH	Cool 4 deg C	PSEG03	E21	02/12/2024	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

02/12/24 16:00
 27600
 CP 5m

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

02/12/24 17:00
 CP 5m
 SO 600,