



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB129497

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/12/2024	16:15
2	CAL2	1	Water	NA	NA	20.3	7.00	02/12/2024	16:16
3	CAL3	1	Water	NA	NA	20.3	10.02	02/12/2024	16:17
4	ICV	1	Water	NA	NA	20.4	7.00	02/12/2024	16:20
5	CCV1	1	Water	NA	NA	20.3	2.01	02/12/2024	16:22
6	P1459-02	1	Solid	20.02	20	23.4	7.99	02/12/2024	16:25
7	P1460-02	1	Solid	20.01	20	23.0	7.75	02/12/2024	16:30
8	P1461-03	1	Solid	20.03	20	23.6	8.14	02/12/2024	16:35
9	P1461-06	1	Solid	20.04	20	23.4	8.05	02/12/2024	16:40
10	P1463-02	1	Solid	20.02	20	23.9	8.10	02/12/2024	16:45
11	P1463-02DUP	1	Solid	20.03	20	24.0	8.11	02/12/2024	16:46
12	P1464-04	1	Solid	20.02	20	24.0	8.78	02/12/2024	16:50
13	P1464-05	1	Solid	20.03	20	24.6	8.51	02/12/2024	17:00
14	P1464-06	1	Solid	20.04	20	24.5	8.49	02/12/2024	17:10
15	CCV2	1	Water	NA	NA	20.3	12.02	02/12/2024	17:15

UP 12/24/24

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity p1459

WorkList ID : 177834

Department : Wet-Chemistry

Date : 02-12-2024 15:48:18

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1459-02	WAL-OILY-DEBRIS	Solid	Corrosivity	Cool 4 deg C	PSEG03	E21	02/12/2024	9045D
P1460-02	MET-OILY-DEBRIS	Solid	Corrosivity	Cool 4 deg C	PSEG03	E21	02/12/2024	9045D
P1461-03	TP-2	Solid	Corrosivity	Cool 4 deg C	PSEG03	E21	02/12/2024	9045D
P1461-06	TP-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	E21	02/12/2024	9045D
P1463-02	SP-G-COMP	Solid	Corrosivity	Cool 4 deg C	PSEG03	E21	02/12/2024	9045D
P1464-04	PN-1(SOUTH-AREA)	Solid	Corrosivity	Cool 4 deg C	EART12	E21	02/12/2024	9045D
P1464-05	PN-2(NORTH-AREA)	Solid	Corrosivity	Cool 4 deg C	SAFE04	E21	02/09/2024	9045D
P1464-06	PN-3(SMALL-STOCK-PILE)	Solid	Corrosivity	Cool 4 deg C	SAFE04	E21	02/09/2024	9045D

Date/Time 02-12-24 16:10

Raw Sample Received by: RJ (EQ-1ab)

Raw Sample Relinquished by:

Date/Time 02-12-24 18:00

Raw Sample Received by: RJ (EQ-1ab)

Raw Sample Relinquished by: RJ (EQ-1ab)