



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129498

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/12/2024	16:15
2	CAL2	1	Water	NA	NA	20.3	7.00	02/12/2024	16:16
3	CAL3	1	Water	NA	NA	20.3	10.02	02/12/2024	16:17
4	ICV	1	Water	NA	NA	20.4	7.00	02/12/2024	16:20
5	CCV1	1	Water	NA	NA	20.3	2.01	02/12/2024	16:22
6	P1463-01	1	Solid	20.02	20	23.9	8.10	02/12/2024	16:45
7	P1463-01DUP	1	Solid	20.03	20	24.0	8.11	02/12/2024	16:46
8	CCV2	1	Water	NA	NA	20.3	12.02	02/12/2024	17:15

WORKLIST(Hardcopy Internal Chain)

VB 129498

WorkList Name : ph p1463

WorkList ID : 177835

Department : Wet-Chemistry

Date : 02-12-2024 15:48:56

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1463-01	SP-G-COMP	Solid	pH	Cool 4 deg C	EART12	E21	02/12/2024	9045D

Date/Time 02-12-24 16:10
 Raw Sample Received by: SD GDCI
 Raw Sample Relinquished by: CP SM

Date/Time 02-12-24 18:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: SD GDCI