

INSTRUMENT ID: CVL

INSTRUMENT ID: CVL



4151E

Linear ▼

A= 0.0000e+000

B= 2.4190e-004

C= -1.0522e-001

Rho= 0.9998925

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	-0.056	-0.056	205	0.000	205	0				0.00
0.2	0.200	0.217	0.017	1334	0.0 %	1334					1.00
2.5	2.500	2.482	-0.018	10694	0.0 %	10694					1.00
5.0	5.000	5.097	0.097	21506	0.0 %	21506					1.00
7.5	7.500	7.522	0.022	31531	0.0 %	31531					1.00
10.0	10.000	9.937	-0.063	41516	0.0 %	41516					1.00

$$\begin{array}{r} 100 \\ - 81 \\ \hline 19 \end{array}$$

LB129504 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Feb 2024 11:25:37

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	205	-	0.0000	7470A	PPB	13 Feb 2024 12:02:22	S	Std
0.2 - 1	50.2	1334	-	0.2000	7470A	PPB	13 Feb 2024 12:04:39	S	Std
2.5 - 1	52.5	10694	-	2.5000	7470A	PPB	13 Feb 2024 12:06:56	S	Std
5.0 - 1	55	21506	-	5.0000	7470A	PPB	13 Feb 2024 12:09:13	S	Std
7.5 - 1	57.5	31531	-	7.5000	7470A	PPB	13 Feb 2024 12:11:31	S	Std
10.0 - 1	59	41516	-	10.0000	7470A	PPB	13 Feb 2024 12:16:34	S	Std
ICV05 - 1	ICV05	17811	4.2032	-	7470A	PPB	13 Feb 2024 12:19:24	U	SMPL
ICB05 - 1	ICB05	225	-0.0508	-	7470A	PPB	13 Feb 2024 12:21:39	U	SMPL
CCV15 - 1	CCV15	21722	5.1493	-	7470A	PPB	13 Feb 2024 12:23:56	U	SMPL
CCB15 - 1	CCB15	225	-0.0508	-	7470A	PPB	13 Feb 2024 12:26:12	U	SMPL
CRA - 1	CRA	1266	0.2010	-	7470A	PPB	13 Feb 2024 12:28:29	U	SMPL
HighStd - 1	HighStd	41710	9.9843	-	7470A	PPB	13 Feb 2024 12:30:45	U	SMPL
ChkStd - 1	ChkStd	30720	7.3259	-	7470A	PPB	13 Feb 2024 12:33:00	U	SMPL
PB158981BL - 1	PBW	127	-0.0745	-	7470A	PPB	13 Feb 2024 12:35:19	U	SMPL
PB158981BS - 1	LCSW	18548	4.3815	-	7470A	PPB	13 Feb 2024 12:37:39	U	SMPL
P1423-02 - 1	TP-3	293	-0.0343	-	7470A	PPB	13 Feb 2024 12:39:55	U	SMPL
P1425-01 - 1	ELZ-24-000012	1565	0.2734	-	7470A	PPB	13 Feb 2024 12:42:14	U	SMPL
P1425-02 - 1	ELZ-24-000006	-51	-0.1176	-	7470A	PPB	13 Feb 2024 12:44:31	U	SMPL
P1425-03 - 1	COMP-1	404	-0.0075	-	7470A	PPB	13 Feb 2024 12:46:48	U	SMPL
P1435-04 - 1	FRENCH-020724	452	0.0041	-	7470A	PPB	13 Feb 2024 12:49:04	U	SMPL
CCV16 - 1	CCV16	21335	5.0557	-	7470A	PPB	13 Feb 2024 12:51:19	U	SMPL
CCB16 - 1	CCB16	175	-0.0629	-	7470A	PPB	13 Feb 2024 12:53:34	U	SMPL
P1446-02 - 1	RT3475	437	0.0005	-	7470A	PPB	13 Feb 2024 12:55:52	U	SMPL
P1449-02 - 1	ETGI-362	459	0.0058	-	7470A	PPB	13 Feb 2024 12:58:08	U	SMPL
P1453-04 - 1	TP-3	432	-0.0007	-	7470A	PPB	13 Feb 2024 13:00:23	U	SMPL
P1453-08 - 1	TP-2	446	0.0027	-	7470A	PPB	13 Feb 2024 13:02:38	U	SMPL
P1453-08DUP - 1	TP-2DUP	543	0.0261	-	7470A	PPB	13 Feb 2024 13:04:54	U	SMPL
P1453-08MS - 1	TP-2MS	17608	4.1541	-	7470A	PPB	13 Feb 2024 13:07:11	U	SMPL
P1453-08MSD - 1	TP-2MSD	17844	4.2112	-	7470A	PPB	13 Feb 2024 13:09:28	U	SMPL
PB158940TB - 1	PB158940TB	237	-0.0479	-	7470A	PPB	13 Feb 2024 13:11:47	U	SMPL
P1453-08LX5 - 1		188	-0.0597	-	7470A	PPB	13 Feb 2024 13:14:06	U	SMPL
P1453-08A - 1		17251	4.0678	-	7470A	PPB	13 Feb 2024 13:16:22	U	SMPL
CCV17 - 1	CCV17	21135	5.0073	-	7470A	PPB	13 Feb 2024 13:18:37	U	SMPL
CCB17 - 1	CCB17	229	-0.0498	-	7470A	PPB	13 Feb 2024 13:20:55	U	SMPL