


LB129508

7471B

INSTRUMENT ID: CV1

 μ Abs.:

6201E



Linear

A= 0.0000e+000

B= 1.6575e-004

C= -5.8818e-002

Rho= 0.9987284

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.000	0.000	355	0.000	355					1/01
0.2	0.200	0.232	0.032	1753	0.0 %	1753					16
2.5	2.500	2.578	0.078	15905	0.0 %	15905					3
5.0	5.000	5.056	0.056	30857	0.0 %	30857					1
7.5	7.500	7.114	-0.386	43275	0.0 %	43275					5
10.0	10.000	10.221	0.221	62016	0.0 %	62016					2

LB129508 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 Feb 2024 13:42:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	58	355	-	0.0000	7471B	PPB	13 Feb 2024 13:54:07	S	Std
0.2 - 1	50.2	1753	-	0.2000	7471B	PPB	13 Feb 2024 13:58:57	S	Std
2.5 - 1	52.5	15905	-	2.5000	7471B	PPB	13 Feb 2024 14:01:14	S	Std
5.0 - 1	55	30857	-	5.0000	7471B	PPB	13 Feb 2024 14:03:32	S	Std
7.5 - 1	57.5	43275	-	7.5000	7471B	PPB	13 Feb 2024 14:05:51	S	Std
10.0 - 1	510	62016	-	10.0000	7471B	PPB	13 Feb 2024 14:08:17	S	Std
ICV06 - 1	ICV06	24570	4.0138	-	7471B	PPB	13 Feb 2024 14:11:21	U	SMPL
ICB06 - 1	ICB06	628	0.0453	-	7471B	PPB	13 Feb 2024 14:13:37	U	SMPL
CCV18 - 1	CCV18	30797	5.0459	-	7471B	PPB	13 Feb 2024 14:16:00	U	SMPL
CCB18 - 1	CCB18	640	0.0473	-	7471B	PPB	13 Feb 2024 14:18:15	U	SMPL
CRA - 1	CRA	1743	0.2301	-	7471B	PPB	13 Feb 2024 14:20:40	U	SMPL
HighStd - 1	HighStd	62265	10.2619	-	7471B	PPB	13 Feb 2024 14:22:55	U	SMPL
ChkStd - 1	ChkStd	40408	6.6390	-	7471B	PPB	13 Feb 2024 14:25:10	U	SMPL
PB158985BL - 1	PBS	313	-0.0069	-	7471B	PPB	13 Feb 2024 14:27:40	U	SMPL
PB158985BS - 1	LCSS	23162	3.7804	-	7471B	PPB	13 Feb 2024 14:30:09	U	SMPL
P1339-01 - 1	WC-COMP	6336	0.9914	-	7471B	PPB	13 Feb 2024 14:32:26	U	SMPL
P1414-01 - 1	SB-6(15-16)	1352	0.1653	-	7471B	PPB	13 Feb 2024 14:34:49	U	SMPL
P1414-01DUP - 1	SB-6(15-16)DUP	1543	0.1969	-	7471B	PPB	13 Feb 2024 14:37:08	U	SMPL
P1414-02 - 1	P1414-01MS	24777	4.0481	-	7471B	PPB	13 Feb 2024 14:39:26	U	SMPL
P1414-03 - 1	P1414-01MSD	26222	4.2876	-	7471B	PPB	13 Feb 2024 14:41:42	U	SMPL
CCV19 - 1	CCV19	30635	5.0191	-	7471B	PPB	13 Feb 2024 14:44:04	U	SMPL
CCB19 - 1	CCB19	683	0.0544	-	7471B	PPB	13 Feb 2024 14:46:26	U	SMPL
P1414-04 - 1	SB-7(12-13)	6164	0.9629	-	7471B	PPB	13 Feb 2024 14:48:51	U	SMPL
P1414-05 - 1	SB-8(10-11)	10327	1.6529	-	7471B	PPB	13 Feb 2024 14:51:06	U	SMPL
P1414-06 - 1	SB-9(10-11)	1201	0.1403	-	7471B	PPB	13 Feb 2024 14:53:24	U	SMPL
P1414-07 - 1	SB-10(10-11)	1268	0.1514	-	7471B	PPB	13 Feb 2024 14:55:42	U	SMPL
P1421-01 - 1	72-11977	26258	4.2936	-	7471B	PPB	13 Feb 2024 14:57:59	U	SMPL
P1423-01 - 1	TP-3	35487	5.8233	-	7471B	PPB	13 Feb 2024 15:00:15	U	SMPL
P1425-04 - 1	VNJ-222	816	0.0764	-	7471B	PPB	13 Feb 2024 15:02:40	U	SMPL
P1448-01 - 1	HD-02-020924	2186	0.3035	-	7471B	PPB	13 Feb 2024 15:05:06	U	SMPL
P1449-01 - 1	ETGI-362	2077	0.2855	-	7471B	PPB	13 Feb 2024 15:07:25	U	SMPL
P1450-01 - 1	OR-02-02092024	2862	0.4156	-	7471B	PPB	13 Feb 2024 15:09:42	U	SMPL
CCV20 - 1	CCV20	31023	5.0834	-	7471B	PPB	13 Feb 2024 15:11:57	U	SMPL
CCB20 - 1	CCB20	731	0.0623	-	7471B	PPB	13 Feb 2024 15:14:14	U	SMPL
P1451-01 - 1	OK-01-02092024	1002	0.1073	-	7471B	PPB	13 Feb 2024 15:16:39	U	SMPL
P1453-01 - 1	TP-3	12768	2.0575	-	7471B	PPB	13 Feb 2024 15:18:54	U	SMPL
P1453-05 - 1	TP-2	8548	1.3581	-	7471B	PPB	13 Feb 2024 15:21:10	U	SMPL
P1414-01LX5 - 1		12	-0.0568	-	7471B	PPB	13 Feb 2024 15:23:28	U	SMPL
P1414-01A - 1		25413	4.1535	-	7471B	PPB	13 Feb 2024 15:25:47	U	SMPL
CCV21 - 1	CCV21	30981	5.0764	-	7471B	PPB	13 Feb 2024 15:28:04	U	SMPL
CCB21 - 1	CCB21	84	-0.0449	-	7471B	PPB	13 Feb 2024 15:30:28	U	SMPL