



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB129510

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP106568
Turbidity Calibration std, 0NTU	WP106567
Turbidity Calibration std, 20NTU	WP106562
Turbidity Calibration std, 40NTU	WP106561
Turbidity Calibration std, 1NTU	WP106565
Turbidity Calibration std, 5NTU	WP106564
Turbidity LOQ std, 1.0NTU	WP106569
Turbidity Calibration - CCV std, 10 NTU	WP106563
10 NTU Standard 500 ml	W2971

Intercept: 0.2465

Slope: 1.0547

Regression: 0.9994

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.156	-0.09		02/13/2024	11:00
2	CAL2	1	1	1.220	0.92	-7.7	02/13/2024	11:03
3	CAL3	5	1	5.360	4.85	-3	02/13/2024	11:06
4	CAL4	10	1	11.700	10.86	8.6	02/13/2024	11:10
5	CAL5	20	1	20.600	19.30	-3.5	02/13/2024	11:13
6	CAL6	40	1	42.600	40.16	0.4	02/13/2024	11:16



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB129510

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	11.000	10.196	02/13/2024	11:20
2	ICB		1	0.160	-0.082	02/13/2024	11:23
3	CCV1	10	1	11.500	10.670	02/13/2024	11:26
4	CCB1		1	0.164	-0.078	02/13/2024	11:30
5	LB129510BL		1	0.160	-0.082	02/13/2024	11:34
6	P1174-02		1	2.760	2.383	02/13/2024	11:37
7	P1174-02DUP		1	2.760	2.383	02/13/2024	11:40
8	P1187-10		1	0.866	0.587	02/13/2024	11:43
9	P1187-11		1	1.420	1.113	02/13/2024	11:47
10	CCV2	10	1	10.600	9.817	02/13/2024	11:50
11	CCB2		1	0.162	-0.080	02/13/2024	11:54