


LB129522

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.7241e-004

C= -9.0136e-002

Rho= 0.9991226

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/11
0.0	0.000	-0.004	-0.004	502	0.000	502					
0.2	0.200	0.232	0.032	1869	0.0 %	1869					16
2.5	2.500	2.367	-0.133	14254	0.0 %	14254					-5
5.0	5.000	5.031	0.031	29701	0.0 %	29701					1
7.5	7.500	7.783	0.283	45665	0.0 %	45665					4
10.0	10.000	9.791	-0.209	57312	0.0 %	57312					-2

LB129522 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Feb 2024 08:45:06

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	502	-	0.0000	7471B	PPB	14 Feb 2024 08:53:26	S	Std
0.2 - 1	50.2	1865	-	0.2000	7471B	PPB	14 Feb 2024 08:55:43	S	Std
2.5 - 1	52.5	14254	-	2.5000	7471B	PPB	14 Feb 2024 08:58:00	S	Std
5.0 - 1	55	29701	-	5.0000	7471B	PPB	14 Feb 2024 09:00:16	S	Std
7.5 - 1	57.5	45665	-	7.5000	7471B	PPB	14 Feb 2024 09:02:36	S	Std
10.0 - 1	510	57312	-	10.0000	7471B	PPB	14 Feb 2024 09:05:01	S	Std
ICV07 - 1	ICV07	24829	4.1905	-	7471B	PPB	14 Feb 2024 09:08:04	U	SMPL
ICB07 - 1	ICB07	112	-0.0708	-	7471B	PPB	14 Feb 2024 09:10:21	U	SMPL
CCV22 - 1	CCV22	29345	4.9691	-	7471B	PPB	14 Feb 2024 09:12:43	U	SMPL
CCB22 - 1	CCB22	101	-0.0727	-	7471B	PPB	14 Feb 2024 09:14:58	U	SMPL
CRA - 1	CRA	1853	0.2293	-	7471B	PPB	14 Feb 2024 09:17:24	U	SMPL
HighStd - 1	HighStd	56809	9.7041	-	7471B	PPB	14 Feb 2024 09:19:39	U	SMPL
ChkStd - 1	ChkStd	43277	7.3711	-	7471B	PPB	14 Feb 2024 09:21:55	U	SMPL
PB158984BL - 1	PBS	372	-0.0260	-	7471B	PPB	14 Feb 2024 09:24:27	U	SMPL
PB158984BS - 1	LCSS	19852	3.3325	-	7471B	PPB	14 Feb 2024 09:26:55	U	SMPL
P1275-01 - 1	TP-2-0.5-0.75	9718	1.5853	-	7471B	PPB	14 Feb 2024 09:29:12	U	SMPL
P1275-02 - 1	TP-3-1-1.5	4796	0.7367	-	7471B	PPB	14 Feb 2024 09:31:37	U	SMPL
P1275-03 - 1	TP-4-1.5-1.75	5361	0.8341	-	7471B	PPB	14 Feb 2024 09:33:56	U	SMPL
P1275-04 - 1	TP-5-0.75-1	14958	2.4887	-	7471B	PPB	14 Feb 2024 09:36:15	U	SMPL
P1275-05 - 1	TP-6-0.5-0.66	2649	0.3666	-	7471B	PPB	14 Feb 2024 09:38:33	U	SMPL
CCV23 - 1	CCV23	28952	4.9014	-	7471B	PPB	14 Feb 2024 09:40:53	U	SMPL
CCB23 - 1	CCB23	68	-0.0784	-	7471B	PPB	14 Feb 2024 09:43:09	U	SMPL
P1275-06 - 1	TP-7-0.5-1.0	5574	0.8709	-	7471B	PPB	14 Feb 2024 09:45:34	U	SMPL
P1275-07 - 1	TP-8-0.75-1.0	1939	0.2442	-	7471B	PPB	14 Feb 2024 09:47:50	U	SMPL
P1394-01 - 1	SB-1(4-5)	953	0.0742	-	7471B	PPB	14 Feb 2024 09:50:08	U	SMPL
P1394-02 - 1	SB-2(3-4)	1115	0.1021	-	7471B	PPB	14 Feb 2024 09:52:25	U	SMPL
P1394-03 - 1	SB-3(4-5)	1015	0.0849	-	7471B	PPB	14 Feb 2024 09:54:41	U	SMPL
P1394-04 - 1	SB-4(5-6)	1351	0.1428	-	7471B	PPB	14 Feb 2024 09:56:57	U	SMPL
P1394-05 - 1	DUP-020624	1020	0.0857	-	7471B	PPB	14 Feb 2024 09:59:14	U	SMPL
P1394-06 - 1	SB-5(13-14)	1740	0.2099	-	7471B	PPB	14 Feb 2024 10:01:31	U	SMPL
P1395-01 - 1	B-4(4-6)	5566	0.8695	-	7471B	PPB	14 Feb 2024 10:03:48	U	SMPL
P1395-01DUP - 1	B-4(4-6)DUP	5342	0.8309	-	7471B	PPB	14 Feb 2024 10:06:04	U	SMPL
CCV24 - 1	CCV24	29740	5.0372	-	7471B	PPB	14 Feb 2024 10:08:22	U	SMPL
CCB24 - 1	CCB24	929	0.0700	-	7471B	PPB	14 Feb 2024 10:10:40	U	SMPL
P1395-01MS - 1	B-4(4-6)MS	30927	5.2419	-	7471B	PPB	14 Feb 2024 10:13:05	U	SMPL
P1395-01MSD - 1	B-4(4-6)MSD	28598	4.8403	-	7471B	PPB	14 Feb 2024 10:15:20	U	SMPL
P1395-02 - 1	B-3(10-12)	8950	1.4529	-	7471B	PPB	14 Feb 2024 10:17:47	U	SMPL
P1395-03 - 1	B-2(4-6)	248231	42.7065	-	7471B	PPB	14 Feb 2024 10:20:14	U	SMPL
P1395-04 - 1	B-2(6-8)	406951	HIGH	-	7471B	PPB	14 Feb 2024 10:22:33	X	
PB159008BL - 1	pbs	581	0.0100	-	7471B	PPB	14 Feb 2024 10:28:02	U	SMPL
PB159008BS - 1	lcss	23488	3.9593	-	7471B	PPB	14 Feb 2024 10:30:17	U	SMPL
P1412-01 - 1	TP2	38557	6.5573	-	7471B	PPB	14 Feb 2024 10:32:34	U	SMPL
P1446-01 - 1	RT3475	210	-0.0539	-	7471B	PPB	14 Feb 2024 10:35:00	U	SMPL
P1454-01 - 1	VNJ-241	2911	0.4117	-	7471B	PPB	14 Feb 2024 10:37:27	U	SMPL
CCV25 - 1	CCV25	30280	5.1303	-	7471B	PPB	14 Feb 2024 10:39:42	U	SMPL
CCB25 - 1	CCB25	860	0.0581	-	7471B	PPB	14 Feb 2024 10:41:59	U	SMPL
P1459-02 - 1	WAL-OILY-DEBRIS	206	-0.0546	-	7471B	PPB	14 Feb 2024 10:44:24	U	SMPL
P1460-02 - 1	MET-OILY-DEBRIS	268	-0.0439	-	7471B	PPB	14 Feb 2024 10:46:40	U	SMPL
P1461-01 - 1	TP-2	20337	3.4161	-	7471B	PPB	14 Feb 2024 10:48:55	U	SMPL
P1461-01DUP - 1	TP-2DUP	21777	3.6644	-	7471B	PPB	14 Feb 2024 10:51:11	U	SMPL
P1461-01MS - 1	TP-2MS	50885	8.6828	-	7471B	PPB	14 Feb 2024 10:53:33	U	SMPL
P1461-01MSD - 1	TP-2MSD	46661	7.9545	-	7471B	PPB	14 Feb 2024 10:55:55	U	SMPL
P1461-04 - 1	TP-1	2023	0.2586	-	7471B	PPB	14 Feb 2024 10:58:24	U	SMPL
P1463-01 - 1	SP-F-COMP	48665	8.3000	-	7471B	PPB	14 Feb 2024 11:00:53	U	SMPL
P1464-01 - 1	PN-1(SOUTH-AREA)	28216	4.7745	-	7471B	PPB	14 Feb 2024 11:03:10	U	SMPL
P1464-02 - 1	PN-2(NORTH-AREA)	17234	2.8811	-	7471B	PPB	14 Feb 2024 11:05:38	U	SMPL
CCV26 - 1	CCV26	30375	5.1467	-	7471B	PPB	14 Feb 2024 11:08:03	U	SMPL
CCB26 - 1	CCB26	873	0.0604	-	7471B	PPB	14 Feb 2024 11:10:22	U	SMPL
P1464-03 - 1	PN-3(SMALL-STOCK-PILE)	24843	4.1930	-	7471B	PPB	14 Feb 2024 11:12:47	U	SMPL
P1395-01LX5 - 1		719	0.0338	-	7471B	PPB	14 Feb 2024 11:15:02	U	SMPL
P1395-01A - 1		29193	4.9429	-	7471B	PPB	14 Feb 2024 11:17:25	U	SMPL
P1461-01LX5 - 1		4807	0.7386	-	7471B	PPB	14 Feb 2024 11:19:40	U	SMPL
P1461-01A - 1		46221	7.8787	-	7471B	PPB	14 Feb 2024 11:22:03	U	SMPL
P1395-03DLX10 - 1		25050	4.2286	-	7471B	PPB	14 Feb 2024 11:24:21	U	SMPL
P1395-04DLX10 - 1		52996	9.0467	-	7471B	PPB	14 Feb 2024 11:26:49	U	SMPL
CCV27 - 1	CCV27	29720	5.0338	-	7471B	PPB	14 Feb 2024 11:29:11	U	SMPL
CCB27 - 1	CCB27	338	-0.0319	-	7471B	PPB	14 Feb 2024 11:31:41	U	SMPL