



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 2/15/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 14:25
In Date: 02/14/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:55
Out Date: 02/15/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB129525

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P1442-01	BH9Q7	1	1.15	8.83	9.98	8.73	85.8	
P1442-02	BH9Q8	2	1.18	8.73	9.91	8.67	85.8	
P1442-03	BH9T6	3	1.19	8.78	9.97	8.86	87.4	
P1442-04	BH9T7	4	1.18	8.63	9.81	8.97	90.3	
P1442-05	VHBLK001	5	1.00	1.00	2.00	2.00	100.0	vhblk
P1442-06	BH9Q9	6	1.13	8.56	9.69	8.62	87.5	
P1442-07	BH9Q9MS	7	1.13	8.56	9.69	8.62	87.5	
P1442-08	BH9Q9MSD	8	1.13	8.56	9.69	8.62	87.5	
P1442-09	BH9T8	9	1.17	8.35	9.52	8.37	86.2	
P1442-10	BH9T9	10	1.19	8.58	9.77	9.6	98.0	
P1442-11	BH9W0	11	1.15	8.78	9.93	8.66	85.5	
P1442-12	BH9W1	12	1.14	8.64	9.78	9.46	96.3	
P1442-13	BH9Q6	13	1.15	8.82	9.97	8.54	83.8	
P1442-14	BH9S0	14	1.19	8.79	9.98	8.72	85.7	
P1442-15	BH9T4	15	1.19	8.55	9.74	8.6	86.7	
P1442-16	BH9T5	16	1.12	8.76	9.88	8.82	87.9	
P1442-17	BH9W2	17	1.19	8.63	9.82	8.66	86.6	
P1442-18	BH9W3	18	1.15	8.44	9.59	8.58	88.0	
P1442-19	BH9R1	19	1.16	8.50	9.66	8.49	86.2	
P1442-20	BH9R2	20	1.14	8.83	9.97	9.05	89.6	
P1442-21	BH9R3	21	1.18	8.79	9.97	9.31	92.5	
P1442-22	BH9W4	22	1.13	8.75	9.88	8.58	85.1	
P1442-23	BH9W5	23	1.16	8.80	9.96	9.26	92.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$