

LB129527


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42945



A= 0.0000e+000

B= 2.3391e-004

C= -1.3774e-002

Rho= 0.9998507

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	u/b D
0.0	0.000	0.010	0.010	103	0.000	103	0				0
0.2	0.200	0.192	-0.008	878	0.0 %	878					-4
2.5	2.500	2.465	-0.035	10595	0.0 %	10595					-1
5.0	5.000	5.106	0.106	21887	0.0 %	21887					2
7.5	7.500	7.396	-0.104	31679	0.0 %	31679					-1
10.0	10.000	10.031	0.031	42945	0.0 %	42945					0

LB129527 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 Feb 2024 13:14:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	103	-	0.0000	7470A	PPB	14 Feb 2024 13:23:31	S	Std
0.2 - 1	SD 2	878	-	0.2000	7470A	PPB	14 Feb 2024 13:25:48	S	Std
2.5 - 1	SD 5	10595	-	2.5000	7470A	PPB	14 Feb 2024 13:28:06	S	Std
5.0 - 1	SD 5	21887	-	5.0000	7470A	PPB	14 Feb 2024 13:30:23	S	Std
7.5 - 1	SD 5	31679	-	7.5000	7470A	PPB	14 Feb 2024 13:32:41	S	Std
10.0 - 1	SD 10	42945	-	10.0000	7470A	PPB	14 Feb 2024 13:34:59	S	Std
ICV08 - 1	ICV08	17152	3.9983	-	7470A	PPB	14 Feb 2024 13:38:21	U	SMPL
ICB08 - 1	ICB08	-178	-0.0554	-	7470A	PPB	14 Feb 2024 13:40:36	U	SMPL
CCV28 - 1	CCV28	22060	5.1463	-	7470A	PPB	14 Feb 2024 13:42:54	U	SMPL
CCB28 - 1	CCB28	-205	-0.0617	-	7470A	PPB	14 Feb 2024 13:45:08	U	SMPL
CRA - 1	CRA	828	0.1799	-	7470A	PPB	14 Feb 2024 13:47:25	U	SMPL
HighStd - 1	HighStd	42593	9.9492	-	7470A	PPB	14 Feb 2024 13:49:41	U	SMPL
ChkStd - 1	ChkStd	29370	6.8562	-	7470A	PPB	14 Feb 2024 13:51:56	U	SMPL
PB159020BL - 1	PBW	-340	-0.0933	-	7470A	PPB	14 Feb 2024 13:54:16	U	SMPL
PB159020BS - 1	LCSW	19042	4.4403	-	7470A	PPB	14 Feb 2024 13:56:34	U	SMPL
P1275-09 - 1	FB	-190	-0.0582	-	7470A	PPB	14 Feb 2024 13:58:50	U	SMPL
P1440-01 - 1	TW-1	3755	0.8646	-	7470A	PPB	14 Feb 2024 14:05:45	U	SMPL
P1440-01DUP - 1	TW-1DUP	3632	0.8358	-	7470A	PPB	14 Feb 2024 14:08:01	U	SMPL
P1440-01MS - 1	TW-1MS	22733	5.3037	-	7470A	PPB	14 Feb 2024 14:10:18	U	SMPL
P1440-01MSD - 1	TW-1MSD	21660	5.0527	-	7470A	PPB	14 Feb 2024 14:12:34	U	SMPL
CCV29 - 1	CCV29	22199	5.1788	-	7470A	PPB	14 Feb 2024 14:14:51	U	SMPL
CCB29 - 1	CCB29	-101	-0.0374	-	7470A	PPB	14 Feb 2024 14:17:09	U	SMPL
P1440-02 - 1	TW-1	35	-0.0056	-	7470A	PPB	14 Feb 2024 14:19:27	U	SMPL
P1440-03 - 1	TW-2	17397	4.0556	-	7470A	PPB	14 Feb 2024 14:21:42	U	SMPL
P1440-04 - 1	TW-2	6	-0.0124	-	7470A	PPB	14 Feb 2024 14:23:58	U	SMPL
P1440-06 - 1	FB020724GW	90	0.0073	-	7470A	PPB	14 Feb 2024 14:26:16	U	SMPL
P1458-01 - 1	A1163	53	-0.0014	-	7470A	PPB	14 Feb 2024 14:28:32	U	SMPL
P1440-01LX5 - 1		746	0.1607	-	7470A	PPB	14 Feb 2024 14:30:48	U	SMPL
P1440-01A - 1		21078	4.9166	-	7470A	PPB	14 Feb 2024 14:33:05	U	SMPL
CCV30 - 1	CCV30	22357	5.2158	-	7470A	PPB	14 Feb 2024 14:38:03	U	SMPL
CCB30 - 1	CCB30	-86	-0.0339	-	7470A	PPB	14 Feb 2024 14:40:19	U	SMPL