



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129528

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	02/14/2024	12:00
2	P1424-04		1	9	55.00	131.0	130.0	02/14/2024	12:30
3	P1424-04DUP		1	9	55.00	131.0	130.0	02/14/2024	13:00
4	P1424-05		1	8	42.00	107.6	106.6	02/14/2024	13:30
5	P1425-01		1	8	40.00	104.0	103.0	02/14/2024	14:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-129528

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 2/14/24 flash point oil WorkList ID : 177869 Department : Wet-Chemistry Date : 02-14-2024 08:50:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1424-04	1315	Solid	Flash Point	Cool 4 deg C	PSEG03	E22	02/08/2024	1010B
P1424-05	1316	Solid	Flash Point	Cool 4 deg C	PSEG03	E22	02/08/2024	1010B
P1425-01	ELZ-24-000012	Solid	Flash Point	Cool 4 deg C	PSEG03	E11	02/08/2024	1010B

Date/Time 02/14/2024 11:50
Raw Sample Received by: N.P. W.C.
Raw Sample Relinquished by: N.P. W.C.

Date/Time 02/14/2024 14:40
Raw Sample Received by: N.P. W.C.
Raw Sample Relinquished by: N.P. W.C.