



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Rubina

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB129530

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP106601
Turbidity Calibration std, 0NTU	WP106600
Turbidity Calibration std, 20NTU	WP106595
Turbidity Calibration std, 40NTU	WP106594
Turbidity Calibration std, 1NTU	WP106598
Turbidity Calibration std, 5NTU	WP106597
Turbidity Calibration - CCV std, 10 NTU	WP106596
10 NTU Standard 500 ml	W2971

Intercept: 0.1419

Slope: 1.0621

Regression: 0.9995

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.152	0.01		02/14/2024	10:35
2	CAL2	1	1	1.190	0.99	-1.3	02/14/2024	10:38
3	CAL3	5	1	5.330	4.89	-2.3	02/14/2024	10:42
4	CAL4	10	1	11.500	10.69	6.9	02/14/2024	10:45
5	CAL5	20	1	20.500	19.17	-4.2	02/14/2024	10:49
6	CAL6	40	1	42.900	40.26	0.6	02/14/2024	10:52



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Rubina

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB129530

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.800	10.035	02/14/2024	10:55
2	ICB		1	0.156	0.013	02/14/2024	10:59
3	CCV1	10	1	11.300	10.506	02/14/2024	11:03
4	CCB1		1	0.160	0.017	02/14/2024	11:06
5	LB129530BL		1	0.156	0.013	02/14/2024	11:10
6	P1187-12		1	0.762	0.584	02/14/2024	11:13
7	CCV2	10	1	11.100	10.317	02/14/2024	11:15
8	CCB2		1	0.162	0.019	02/14/2024	11:18