

LB129531


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

55117



A= 0.0000e+000

B= 1.8258e-004 *slope*C= 4.9821e-003 *x-intercept*

Rho= 0.9999240

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.00	0.000	0.026	0.026	116	0.000	116					160
0.05	0.050					442	0				200
0.20	0.200	0.204	0.004	1092	0.0 %	1092					200
2.50	2.500	2.499	-0.001	13658	0.0 %	13658					200
5.00	5.000	4.985	-0.015	27274	0.0 %	27274					200
7.50	7.500	7.418	-0.082	40598	0.0 %	40598					200
10.0	10.000	10.068	0.068	55117	0.0 %	55117					200

LB129531

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	116	-		0 SFAM01.1	PPB	2/14/2024 15:21	Std
	0.2 S01	1092	-		0.2 SFAM01.1	PPB	2/14/2024 15:23	Std
	2.5 S02	13658	-		2.5 SFAM01.1	PPB	2/14/2024 15:25	Std
	5 S03	27274	-		5 SFAM01.1	PPB	2/14/2024 15:28	Std
	7.5 S04	40598	-		7.5 SFAM01.1	PPB	2/14/2024 15:30	Std
	10 S05	55117	-		10 SFAM01.1	PPB	2/14/2024 15:32	Std
ICV009	ICV009	22399	4.0947 -		SFAM01.1	PPB	2/14/2024 15:35	SMPL
ICB009	ICB009	-86	-0.0107 -		SFAM01.1	PPB	2/14/2024 15:37	SMPL
CCV031	CCV031	27332	4.9954 -		SFAM01.1	PPB	2/14/2024 15:40	SMPL
CCB031	CCB031	5	0.0059 -		SFAM01.1	PPB	2/14/2024 15:44	SMPL
PB159024BL	PBS024	24	0.0094 -		SFAM01.1	PPB	2/14/2024 15:47	SMPL
P1266-01	MJPTA0	1738	0.3223 -		SFAM01.1	PPB	2/14/2024 15:49	SMPL
P1266-02	MJPTA1	740	0.1401 -		SFAM01.1	PPB	2/14/2024 15:56	SMPL
P1266-03	MJPTA1D	763	0.1443 -		SFAM01.1	PPB	2/14/2024 15:58	SMPL
P1266-04	MJPTA1S	13438	2.4585 -		SFAM01.1	PPB	2/14/2024 16:00	SMPL
CCV032	CCV032	26848	4.907 -		SFAM01.1	PPB	2/14/2024 16:03	SMPL
CCB032	CCB032	-85	-0.0105 -		SFAM01.1	PPB	2/14/2024 16:05	SMPL