

LB129535


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

49138



A= 0.0000e+000

B= 2.0361e-004 slope

C= -2.0059e-002 y-intercept

Rho= 0.9999863

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/1/1)
0.00	0.000	0.016	0.016	176	0.000	176					=
0.05	0.050					442	0				-6
0.20	0.200	0.189	-0.011	1026	0.0 %	1026					-1
2.50	2.500	2.472	-0.028	12240	0.0 %	12240					1
5.00	5.000	5.027	0.027	24787	0.0 %	24787					0
7.50	7.500	7.512	0.012	36992	0.0 %	36992					0
10.0	10.000	9.985	-0.015	49138	0.0 %	49138					0

LB129535

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0	176	-		0 SFAM01.1	PPB	2/14/2024 16:21	Std
	0.2 S01	1026	-		0.2 SFAM01.1	PPB	2/14/2024 16:24	Std
	2.5 S02	12240	-		2.5 SFAM01.1	PPB	2/14/2024 16:26	Std
	5 S03	24787	-		5 SFAM01.1	PPB	2/14/2024 16:28	Std
	7.5 S04	36992	-		7.5 SFAM01.1	PPB	2/14/2024 16:31	Std
	10 S05	49138	-		10 SFAM01.1	PPB	2/14/2024 16:33	Std
ICV010	ICV010	19902	4.0321 -		SFAM01.1	PPB	2/14/2024 16:36	SMPL
ICB010	ICB010	109	0.0021 -		SFAM01.1	PPB	2/14/2024 16:41	SMPL
CCV033	CCV033	24793	5.028 -		SFAM01.1	PPB	2/14/2024 16:43	SMPL
CCB033	CCB033	97	-0.0003 -		SFAM01.1	PPB	2/14/2024 16:48	SMPL
PBW001	PBW001	154	0.0113 -		SFAM01.1	PPB	2/14/2024 16:50	SMPL
PBW002	PBW002	152	0.0109 -		SFAM01.1	PPB	2/14/2024 16:52	SMPL
PBW003	PBW003	157	0.0119 -		SFAM01.1	PPB	2/14/2024 16:54	SMPL
MDL01	MDL01	333	0.0477 -		SFAM01.1	PPB	2/14/2024 16:57	SMPL
MDL02	MDL02	357	0.0526 -		SFAM01.1	PPB	2/14/2024 16:59	SMPL
MDL03	MDL03	341	0.0494 -		SFAM01.1	PPB	2/14/2024 17:01	SMPL
CCV034	CCV034	24779	5.0251 -		SFAM01.1	PPB	2/14/2024 17:04	SMPL
CCB034	CCB034	124	0.0052 -		SFAM01.1	PPB	2/14/2024 17:08	SMPL