



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129539

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/15/2024	08:00
2	CAL2	1	Water	NA	NA	20.2	7.00	02/15/2024	08:01
3	CAL3	1	Water	NA	NA	20.2	10.02	02/15/2024	08:05
4	ICV	1	Water	NA	NA	20.2	7.02	02/15/2024	08:10
5	CCV1	1	Water	NA	NA	20.1	2.02	02/15/2024	08:45
6	P1468-02	1	Water	NA	NA	20.1	5.04	02/15/2024	08:48
7	P1471-04	1	Water	NA	NA	20.5	7.78	02/15/2024	08:50
8	P1482-01	1	Water	NA	NA	20.1	7.57	02/15/2024	09:00
9	P1482-01DUP	1	Water	NA	NA	20.2	7.58	02/15/2024	09:01
10	CCV2	1	Water	NA	NA	20.3	12.02	02/15/2024	09:05

WORKLIST(Hardcopy Internal Chain)

WD 124539

WorkList Name : ph p1482 WorkList ID : 177908 Department : Wet-Chemistry Date : 02-15-2024 07:42:42

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1468-02	A-B-C-CELL-POOL-1ST-Q	Water	pH	Cool 4 deg C	STER12	E21	02/12/2024	9040C
P1471-04	30985	Water	pH	Cool 4 deg C	PSEG03	E21	02/14/2024	9040C
P1482-01	NWB-1991	Water	pH	Cool 4 deg C	PSEG03	E21	02/14/2024	9040C

Date/Time 02-15-24 07:50
 Raw Sample Received by: SOLOCI
 Raw Sample Relinquished by: CL SN

Date/Time 02-15-24 12:00
 Raw Sample Received by: CL SN
 Raw Sample Relinquished by: JO LEE