



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB129540

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/15/2024	08:00
2	CAL2	1	Water	NA	NA	20.2	7.00	02/15/2024	08:01
3	CAL3	1	Water	NA	NA	20.2	10.02	02/15/2024	08:05
4	ICV	1	Water	NA	NA	20.3	7.02	02/15/2024	08:10
5	CCV1	1	Water	NA	NA	20.2	2.01	02/15/2024	09:15
6	P1486-02	1	Solid	20.02	20	23.6	8.24	02/15/2024	09:25
7	P1486-02DUP	1	Solid	20.01	20	23.7	8.25	02/15/2024	09:26
8	P1486-03	1	Solid	20.02	20	24.2	8.59	02/15/2024	09:30
9	CCV2	1	Water	NA	NA	20.3	12.02	02/15/2024	09:35

W0129540

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity p1471 WorkList ID : 177910 Department : Wet-Chemistry Date : 02-15-2024 07:45:54

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1486-02	3420	Solid	Corrosivity	Cool 4 deg C	PSEG03	E21	02/14/2024	9045D
P1486-03	3418-3419	Solid	Corrosivity	Cool 4 deg C	PSEG03	E21	02/14/2024	9045D

Date/Time 02-15-24 07:50
Raw Sample Received by: JD Park
Raw Sample Relinquished by: CP Sr

Date/Time 02-15-24 12:00
Raw Sample Received by: CP Sr
Raw Sample Relinquished by: JD Park