

L3129544


7471B

INSTRUMENT ID: N1

Linear

μ Abs.:

57478



A= 0.0000e+000

B= 1.7447e-004

C= -2.4138e-002

Rho= 0.9998771

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	-0.016	-0.016	46	0.000	46					11.1
0.2	0.200	0.174	-0.026	1138	0.0 %	1138					-13
2.5	2.500	2.610	0.110	15095	0.0 %	15095					4
5.0	5.000	4.916	-0.084	28316	0.0 %	28316					2
7.5	7.500	7.512	0.012	43191	0.0 %	43191					0
10.0	10.000	10.004	0.004	57478	0.0 %	57478					0

LB129544 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 15 Feb 2024 13:36:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	46	-	0.0000	7471B	PPB	15 Feb 2024 13:43:25	S	Std
0.2 - 1	50.2	1138	-	0.2000	7471B	PPB	15 Feb 2024 13:45:43	S	Std
2.5 - 1	52.5	15095	-	2.5000	7471B	PPB	15 Feb 2024 13:48:00	S	Std
5.0 - 1	57.5	28316	-	5.0000	7471B	PPB	15 Feb 2024 13:50:16	S	Std
7.5 - 1	57.5	43191	-	7.5000	7471B	PPB	15 Feb 2024 13:52:36	S	Std
10.0 - 1	57.5	57478	-	10.0000	7471B	PPB	15 Feb 2024 13:57:47	S	Std
ICV11 - 1	ICV11	23163	4.0172	-	7471B	PPB	15 Feb 2024 14:00:42	U	SMPL
ICB11 - 1	ICB11	-278	-0.0726	-	7471B	PPB	15 Feb 2024 14:02:58	U	SMPL
CCV35 - 1	CCV35	28425	4.9353	-	7471B	PPB	15 Feb 2024 14:05:21	U	SMPL
CCB35 - 1	CCB35	-272	-0.0716	-	7471B	PPB	15 Feb 2024 14:07:36	U	SMPL
CRA - 1	CRA	1059	0.1606	-	7471B	PPB	15 Feb 2024 14:10:02	U	SMPL
HighStd - 1	HighStd	57790	10.0587	-	7471B	PPB	15 Feb 2024 14:12:17	U	SMPL
ChkStd - 1	ChkStd	40666	7.0710	-	7471B	PPB	15 Feb 2024 14:14:33	U	SMPL
PB159044BL - 1	PBS	-423	-0.0979	-	7471B	PPB	15 Feb 2024 14:17:04	U	SMPL
PB159044BS - 1	LCSS	23128	4.0111	-	7471B	PPB	15 Feb 2024 14:19:31	U	SMPL
P1435-06 - 1	FRENCH-020724-CONC	2735	0.4530	-	7471B	PPB	15 Feb 2024 14:21:47	U	SMPL
P1472-01 - 1	TR-04-021424	3946	0.6643	-	7471B	PPB	15 Feb 2024 14:30:09	U	SMPL
P1472-01DUP - 1	TR-04-021424DUP	3789	0.6369	-	7471B	PPB	15 Feb 2024 14:32:25	U	SMPL
P1472-01MS - 1	TR-04-021424MS	28369	4.9255	-	7471B	PPB	15 Feb 2024 14:34:44	U	SMPL
P1472-01MSD - 1	TR-04-021424MSD	27422	4.7603	-	7471B	PPB	15 Feb 2024 14:37:02	U	SMPL
CCV36 - 1	CCV36	28765	4.9946	-	7471B	PPB	15 Feb 2024 14:39:22	U	SMPL
CCB36 - 1	CCB36	-253	-0.0683	-	7471B	PPB	15 Feb 2024 14:41:50	U	SMPL
P1473-01 - 1	SU-03-021424	1736	0.2787	-	7471B	PPB	15 Feb 2024 14:44:14	U	SMPL
P1473-03 - 1	SU-04-021424	2771	0.4593	-	7471B	PPB	15 Feb 2024 14:46:30	U	SMPL
P1483-01 - 1	RT-2327	31404	5.4551	-	7471B	PPB	15 Feb 2024 14:48:45	U	SMPL
P1483-03 - 1	VNJ-205	10584	1.8225	-	7471B	PPB	15 Feb 2024 14:51:02	U	SMPL
P1472-01LX5 - 1		536	0.0694	-	7471B	PPB	15 Feb 2024 14:53:24	U	SMPL
P1472-01A - 1		27736	4.8151	-	7471B	PPB	15 Feb 2024 14:55:51	U	SMPL
CCV37 - 1	CCV37	29325	5.0923	-	7471B	PPB	15 Feb 2024 14:58:07	U	SMPL
CCB37 - 1	CCB37	-76	-0.0374	-	7471B	PPB	15 Feb 2024 15:00:27	U	SMPL