



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129553

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	02/15/2024	10:10
2	P1458-01		1	14	100.00	>212.0	>212.0	02/15/2024	10:40
3	P1458-01DUP		1	13	100.00	>212.0	>212.0	02/15/2024	11:10
4	P1471-04		1	14	100.00	>212.0	>212.0	02/15/2024	11:40
5	P1482-01		1	12	100.00	>212.0	>212.0	02/15/2024	12:10
6	P1484-01		1	13	100.00	>212.0	>212.0	02/15/2024	12:40
7	P1484-02		1	12	100.00	>212.0	>212.0	02/15/2024	13:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

CB-129553

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 02/15/24 flashpoint w WorkList ID : 177928 Department : Wet-Chemistry Date : 02-15-2024 08:25:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1458-01	A1163	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/12/2024	1010B
P1471-04	30985	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/14/2024	1010B
P1482-01	NWB-1991	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/14/2024	1010B
P1484-01	382	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/14/2024	1010B
P1484-02	386	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/14/2024	1010B

Date/Time 02/15/24 10:00
Raw Sample Received by: N.P. W.C.
Raw Sample Relinquished by: M. Lacey

Date/Time 02/15/24 13:50
Raw Sample Received by: M. Lacey
Raw Sample Relinquished by: N.P. W.C.