

LB129559


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

52153



A= 0.0000e+000

B= 1.9214e-004 slope

C= 3.4486e-002 y-intercept

Rho= 0.9999287

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.048	0.048	69	0.000	69					
0.05	0.050					442	0				
0.20	0.200	0.224	0.024	985	0.0 %	985					
2.50	2.500	2.449	-0.051	12564	0.0 %	12564					
5.00	5.000	4.950	-0.050	25584	0.0 %	25584					
7.50	7.500	7.475	-0.025	38723	0.0 %	38723					
10.0	10.000	10.055	0.055	52153	0.0 %	52153					

12
12
11
01
1

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	69	-		0 SFAM01.1	PPB	2/16/2024 9:56	Std
	0.2 S01	985	-		0.2 SFAM01.1	PPB	2/16/2024 9:59	Std
	2.5 S02	12564	-		2.5 SFAM01.1	PPB	2/16/2024 10:01	Std
	5 S03	25584	-		5 SFAM01.1	PPB	2/16/2024 10:03	Std
	7.5 S04	38723	-		7.5 SFAM01.1	PPB	2/16/2024 10:05	Std
	10 S05	52153	-		10 SFAM01.1	PPB	2/16/2024 10:08	Std
ICV012	ICV012	21160	4.1002	-	SFAM01.1	PPB	2/16/2024 10:11	SMPL
ICB012	ICB012	-160	0.0037	-	SFAM01.1	PPB	2/16/2024 10:13	SMPL
CCV038	CCV038	25966	5.0236	-	SFAM01.1	PPB	2/16/2024 10:15	SMPL
CCB038	CCB038	-175	0.0009	-	SFAM01.1	PPB	2/16/2024 10:17	SMPL
PB159050BL	PBS050	16	0.0376	-	SFAM01.1	PPB	2/16/2024 10:20	SMPL
P1287-01	MBH6S2	1991	0.417	-	SFAM01.1	PPB	2/16/2024 10:22	SMPL
P1287-02	MBH6S3	601	0.15	-	SFAM01.1	PPB	2/16/2024 10:24	SMPL
P1287-03	MBH6S4	572	0.1444	-	SFAM01.1	PPB	2/16/2024 10:26	SMPL
P1287-04	MBH6T4	408	0.1129	-	SFAM01.1	PPB	2/16/2024 10:29	SMPL
P1287-05	MBH6T5	435	0.1181	-	SFAM01.1	PPB	2/16/2024 10:31	SMPL
P1287-06	MBH6T6	479	0.1265	-	SFAM01.1	PPB	2/16/2024 10:33	SMPL
P1287-07	MBH6W9	11121	2.1713	-	SFAM01.1	PPB	2/16/2024 10:36	SMPL
P1287-08	MBH6X0	514	0.1332	-	SFAM01.1	PPB	2/16/2024 10:38	SMPL
P1287-09	MBH6X1	465	0.1238	-	SFAM01.1	PPB	2/16/2024 10:40	SMPL
P1287-10	MBH6X5	13086	2.5488	-	SFAM01.1	PPB	2/16/2024 10:42	SMPL
P1287-11	MBH6S5	1510	0.3246	-	SFAM01.1	PPB	2/16/2024 10:45	SMPL
P1287-12	MBH6S5D	1305	0.2852	-	SFAM01.1	PPB	2/16/2024 10:47	SMPL
P1287-13	MBH6S5S	14419	2.8049	-	SFAM01.1	PPB	2/16/2024 10:49	SMPL
P1287-14	MBH6S6	230	0.0787	-	SFAM01.1	PPB	2/16/2024 10:51	SMPL
P1287-15	MBH6S7	400	0.1113	-	SFAM01.1	PPB	2/16/2024 10:54	SMPL
P1287-16	MBH6W0	1804	0.3811	-	SFAM01.1	PPB	2/16/2024 10:56	SMPL
P1287-17	MBH6W1	189	0.0708	-	SFAM01.1	PPB	2/16/2024 10:58	SMPL
P1287-18	MBH6W2	421	0.1154	-	SFAM01.1	PPB	2/16/2024 11:01	SMPL
P1287-20	MBH6S8	8695	1.7051	-	SFAM01.1	PPB	2/16/2024 11:03	SMPL
CCV039	CCV039	26985	5.2194	-	SFAM01.1	PPB	2/16/2024 11:05	SMPL
CCB039	CCB039	-98	0.0157	-	SFAM01.1	PPB	2/16/2024 11:07	SMPL
P1287-21	MBH6S9	458	0.1225	-	SFAM01.1	PPB	2/16/2024 11:10	SMPL
P1287-22	MBH6T0	458	0.1225	-	SFAM01.1	PPB	2/16/2024 11:12	SMPL
PB159051BL	PBS051	19	0.0381	-	SFAM01.1	PPB	2/16/2024 11:14	SMPL
P1302-01	MCOFG5	850	0.1978	-	SFAM01.1	PPB	2/16/2024 11:16	SMPL
P1302-02	MCOFG7	1415	0.3064	-	SFAM01.1	PPB	2/16/2024 11:19	SMPL
P1302-03	MCOFH2	1160	0.2574	-	SFAM01.1	PPB	2/16/2024 11:21	SMPL
P1302-04	MCOFH3	2680	0.5494	-	SFAM01.1	PPB	2/16/2024 11:23	SMPL
P1302-05	MCOFH3D	2413	0.4981	-	SFAM01.1	PPB	2/16/2024 11:25	SMPL
P1302-06	MCOFH3S	16451	3.1954	-	SFAM01.1	PPB	2/16/2024 11:28	SMPL
P1302-07	MCOFH4	822	0.1924	-	SFAM01.1	PPB	2/16/2024 11:30	SMPL
P1302-08	MCOFH8	798	0.1878	-	SFAM01.1	PPB	2/16/2024 11:32	SMPL
P1302-09	MCOFH9	2967	0.6046	-	SFAM01.1	PPB	2/16/2024 11:35	SMPL
P1302-10	MCOFJ0	1096	0.2451	-	SFAM01.1	PPB	2/16/2024 11:37	SMPL
P1302-11	MCOFJ1	1698	0.3607	-	SFAM01.1	PPB	2/16/2024 11:39	SMPL

INSTRUMENT ID : CV1

P1302-12	MCOFJ2	3267	0.6622 -	SFAM01.1	PPB	2/16/2024 11:41 SMPL
P1302-13	MCOFJ3	1349	0.2937 -	SFAM01.1	PPB	2/16/2024 11:44 SMPL
P1302-14	MCOFH5	1128	0.2512 -	SFAM01.1	PPB	2/16/2024 11:46 SMPL
P1302-15	MCOFH6	1654	0.3523 -	SFAM01.1	PPB	2/16/2024 11:48 SMPL
P1302-16	MCOFJ4	1204	0.2658 -	SFAM01.1	PPB	2/16/2024 11:50 SMPL
P1302-17	MCOFJ5	1657	0.3529 -	SFAM01.1	PPB	2/16/2024 11:53 SMPL
CCV040	CCV040	26704	5.1654 -	SFAM01.1	PPB	2/16/2024 11:55 SMPL
CCB040	CCB040	-90	0.0172 -	SFAM01.1	PPB	2/16/2024 11:57 SMPL
P1302-18	MCOFJ7	3336	0.6755 -	SFAM01.1	PPB	2/16/2024 12:00 SMPL
P1302-19	MCOFK0	10766	2.1031 -	SFAM01.1	PPB	2/16/2024 12:02 SMPL
P1302-20	MCOFK3	5493	1.0899 -	SFAM01.1	PPB	2/16/2024 12:04 SMPL
P1302-21	MCOFK9	1117	0.2491 -	SFAM01.1	PPB	2/16/2024 12:06 SMPL
P1302-22	MCOFLO	3188	0.647 -	SFAM01.1	PPB	2/16/2024 12:09 SMPL
PB159052BL	PBS052	2	0.0349 -	SFAM01.1	PPB	2/16/2024 12:11 SMPL
P1328-01	MCOFK8	1555	0.3333 -	SFAM01.1	PPB	2/16/2024 12:13 SMPL
P1328-02	MCOFK8D	1450	0.3131 -	SFAM01.1	PPB	2/16/2024 12:15 SMPL
P1328-03	MCOFK8S	15138	2.9431 -	SFAM01.1	PPB	2/16/2024 12:18 SMPL
P1328-04	MCOFL2	932	0.2136 -	SFAM01.1	PPB	2/16/2024 12:20 SMPL
P1328-05	MCOFL3	1430	0.3092 -	SFAM01.1	PPB	2/16/2024 12:22 SMPL
P1328-06	MCOFL5	2857	0.5834 -	SFAM01.1	PPB	2/16/2024 12:25 SMPL
P1328-07	MCOFL6	3296	0.6678 -	SFAM01.1	PPB	2/16/2024 12:27 SMPL
P1328-08	MCOFL7	2273	0.4712 -	SFAM01.1	PPB	2/16/2024 12:29 SMPL
P1328-09	MCOFL8	2017	0.422 -	SFAM01.1	PPB	2/16/2024 12:31 SMPL
P1328-10	MCOFM1	2021	0.4228 -	SFAM01.1	PPB	2/16/2024 12:34 SMPL
P1328-11	MCOFM2	2989	0.6088 -	SFAM01.1	PPB	2/16/2024 12:36 SMPL
P1328-12	MCOFM7	1895	0.3986 -	SFAM01.1	PPB	2/16/2024 12:38 SMPL
P1328-13	MCOFR5	1144	0.2543 -	SFAM01.1	PPB	2/16/2024 12:41 SMPL
P1328-14	MCOFR6	8499	1.6675 -	SFAM01.1	PPB	2/16/2024 12:43 SMPL
CCV041	CCV041	26196	5.0678 -	SFAM01.1	PPB	2/16/2024 12:45 SMPL
CCB041	CCB041	-35	0.0278 -	SFAM01.1	PPB	2/16/2024 12:47 SMPL
P1328-15	MCOFR7	3071	0.6245 -	SFAM01.1	PPB	2/16/2024 12:50 SMPL
P1328-16	MCOFM4	1547	0.3317 -	SFAM01.1	PPB	2/16/2024 12:52 SMPL
P1328-17	MCOFM5	1964	0.4118 -	SFAM01.1	PPB	2/16/2024 12:54 SMPL
P1328-18	MCOFM6	1387	0.301 -	SFAM01.1	PPB	2/16/2024 12:56 SMPL
P1328-19	MCOFNO	1938	0.4069 -	SFAM01.1	PPB	2/16/2024 12:59 SMPL
P1328-20	MCOFN1	1294	0.2831 -	SFAM01.1	PPB	2/16/2024 13:01 SMPL
P1328-21	MCOFN6	3398	0.6874 -	SFAM01.1	PPB	2/16/2024 13:03 SMPL
P1328-22	MCOFN7	4709	0.9393 -	SFAM01.1	PPB	2/16/2024 13:06 SMPL
CCV042	CCV042	25754	4.9828 -	SFAM01.1	PPB	2/16/2024 13:08 SMPL
CCB042	CCB042	-86	0.018 -	SFAM01.1	PPB	2/16/2024 13:10 SMPL