



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129562

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	02/16/2024	12:30
2	P1486-04		1	12	100.00	>212.0	>212.0	02/16/2024	13:00
3	P1486-04DUP		1	11	100.00	>212.0	>212.0	02/16/2024	13:30
4	P1486-05		1	14	100.00	>212.0	>212.0	02/16/2024	14:00
5	P1486-06		1	13	100.00	>212.0	>212.0	02/16/2024	14:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-129562

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 02/16/24 flashpoint w WorkList ID : 177937 Department : Wet-Chemistry Date : 02-16-2024 08:11:58

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1486-04	3397	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/14/2024	1010B
P1486-05	3428	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/14/2024	1010B
P1486-06	3417	Water	Flash Point	Cool 4 deg C	PSEG03	E21	02/14/2024	1010B

Date/Time 02/16/2024 12:20
Raw Sample Received by: N.P. W.C.
Raw Sample Relinquished by: N.P. W.C.

Date/Time 02/16/2024 14:40
Raw Sample Received by: N.P. W.C.
Raw Sample Relinquished by: N.P. W.C.