

Analysis Method: 9034

Parameter: Reactive Sulfide

Run Number: LB129566

ANALYST: Nikita

SUPERVISOR REVIEW BY: Iwona

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W2976
IODINE SOLUTION .025N 1L	W3050
Starch Solution, 4L	W2977

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB159073BL		1	5.00	50	2.00	0.00	1.92	1.92	0.08	0.00	0.00	02/16/2024	11:15
2	P1486-01		1	5.02	50	2.00	0.00	1.90	1.90	0.10	0.02	1.59	02/16/2024	11:18
3	P1486-01DUP		1	5.04	50	2.00	0.00	1.90	1.90	0.10	0.02	1.59	02/16/2024	11:21
4	P1486-02		1	5.03	50	2.00	0.00	1.88	1.88	0.12	0.04	3.18	02/16/2024	11:24
5	P1486-03		1	5.06	50	2.00	0.00	1.86	1.86	0.14	0.06	4.74	02/16/2024	11:27
6	P1488-01		1	5.04	50	2.00	0.00	1.90	1.90	0.10	0.02	1.59	02/16/2024	11:30
7	P1505-04		1	5.07	50	2.00	0.00	1.88	1.88	0.12	0.04	3.16	02/16/2024	11:33
8	P1505-08		1	5.05	50	2.00	0.00	1.86	1.86	0.14	0.06	4.75	02/16/2024	11:36
9	P1505-12		1	5.02	50	2.00	0.00	1.86	1.86	0.14	0.06	4.78	02/16/2024	11:39
10	P1505-16		1	5.01	50	2.00	0.00	1.88	1.88	0.12	0.04	3.19	02/16/2024	11:42

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume