


LB129569

SFAM01.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000
 B= 2.1095e-004 *slope*
 C= -3.0394e-002 *= intercept*
 Rho= 0.9999681
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>o/e D</i>
0.00	0.000	0.000	0.000	145	0.000	145					
0.05	0.050					442	0				
0.20	0.200	0.177	-0.023	983	0.0 %	983					
2.50	2.500	2.507	0.007	12027	0.0 %	12027					
5.00	5.000	5.011	0.011	23896	0.0 %	23896					
7.50	7.500	7.550	0.050	35935	0.0 %	35935					
10.0	10.000	9.955	-0.045	47336	0.0 %	47336					

o/e D
 =
 -12
 0
 0
 1
 0

LB129569

INSTRUMENT ID : 1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	145	-		0 SFAM01.1	PPB	2/16/2024 13:40	Std
	0.2 S01	983	-		0.2 SFAM01.1	PPB	2/16/2024 13:42	Std
	2.5 S02	12027	-		2.5 SFAM01.1	PPB	2/16/2024 13:44	Std
	5 S03	23896	-		5 SFAM01.1	PPB	2/16/2024 13:47	Std
	7.5 S04	35935	-		7.5 SFAM01.1	PPB	2/16/2024 13:49	Std
	10 S05	47336	-		10 SFAM01.1	PPB	2/16/2024 13:51	Std
ICV013	ICV013	19539	4.0914 -		SFAM01.1	PPB	2/16/2024 13:54	SMPL
ICB013	ICB013	111	-0.007 -		SFAM01.1	PPB	2/16/2024 13:59	SMPL
CCV043	CCV043	24123	5.0584 -		SFAM01.1	PPB	2/16/2024 14:01	SMPL
CCB043	CCB043	112	-0.0068 -		SFAM01.1	PPB	2/16/2024 14:06	SMPL
PBW06	PBW06	120	-0.0051 -		SFAM01.1	PPB	2/16/2024 14:08	SMPL
PBW07	PBW07	110	-0.0072 -		SFAM01.1	PPB	2/16/2024 14:10	SMPL
MDL06	MDL06	316	0.0363 -		SFAM01.1	PPB	2/16/2024 14:13	SMPL
MDL07	MDL07	319	0.0369 -		SFAM01.1	PPB	2/16/2024 14:15	SMPL
CCV044	CCV044	24058	5.0447 -		SFAM01.1	PPB	2/16/2024 14:17	SMPL
CCB044	CCB044	110	-0.0072 -		SFAM01.1	PPB	2/16/2024 14:22	SMPL