



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 2/19/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 15:25
In Date: 02/16/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: m oven#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:55
Out Date: 02/17/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: m sc-4
Thermometer ID: % solids-oven

QC:LB129571

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P1470-01	MBH9R0	1	1.18	8.42	9.6	8.67	89.0	
P1470-02	MBH9R0D	2	1.18	8.42	9.6	8.67	89.0	
P1470-03	MBH9R0S	3	1.18	8.42	9.6	8.67	89.0	
P1470-04	MBH9W4	4	1.13	8.75	9.88	8.58	85.1	
P1470-05	MBH9W5	5	1.16	8.80	9.96	9.26	92.0	
P1470-06	MBH9W6	6	1.18	8.59	9.77	8.73	87.9	
P1470-07	MBH9W7	7	1.19	8.62	9.81	8.54	85.3	
P1470-08	MBH9R5	8	1.19	8.50	9.69	8.39	84.7	
P1470-09	MBH9R6	9	1.19	8.51	9.7	8.81	89.5	
P1470-10	MBH9X0	10	1.15	8.82	9.97	8.76	86.3	
P1470-11	MBH9X2	11	1.16	8.69	9.85	8.25	81.6	
P1470-12	MBH9X3	12	1.19	8.52	9.71	7.99	79.8	
P1470-13	MBH9X4	13	1.15	8.50	9.65	8.63	88.0	
P1470-14	MBH9X5	14	1.14	8.48	9.62	8.66	88.7	
P1470-15	MBH9R4	15	1.15	8.82	9.97	8.79	86.6	
P1470-16	MBH9Q5	16	1.14	8.84	9.98	7.42	71.0	
P1470-17	MBH9R7	17	1.16	8.79	9.95	8.94	88.5	
P1470-18	MBH9R8	18	1.14	8.83	9.97	8.48	83.1	
P1470-19	MBH9S1	19	1.19	8.38	9.57	8.09	82.3	
P1470-20	MBH9X6	20	1.19	8.64	9.83	8.45	84.0	
P1470-21	MBH9X7	21	1.13	8.44	9.57	8.36	85.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$