


LB 129575

7470A

INSTRUMENT ID: CV1

 μ Abs.:

4152E



Linear

A= 0.0000e+000

B= 2.3821e-004

C= -1.6653e-002

Rho= 0.9995021

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	W/D
0.0	0.000	0.035	0.035	217	0.000	217	0				-
0.2	0.200	0.222	0.022	1002	0.0 %	1002					11
2.5	2.500	2.395	-0.105	10122	0.0 %	10122					-4
5.0	5.000	4.949	-0.051	20846	0.0 %	20846					-1
7.5	7.500	7.724	0.224	32496	0.0 %	32496					3
10.0	10.000	9.875	-0.125	41525	0.0 %	41525					-1

LB129575 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Feb 2024 14:44:26

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	217	-	0.0000	7470A	PPB	16 Feb 2024 14:46:07	S	Std
0.2 - 1	50.2	1002	-	0.2000	7470A	PPB	16 Feb 2024 14:48:25	S	Std
2.5 - 1	52.5	10122	-	2.5000	7470A	PPB	16 Feb 2024 14:50:42	S	Std
5.0 - 1	55	20846	-	5.0000	7470A	PPB	16 Feb 2024 14:52:58	S	Std
7.5 - 1	57.5	32496	-	7.5000	7470A	PPB	16 Feb 2024 14:55:17	S	Std
10.0 - 1	50	41525	-	10.0000	7470A	PPB	16 Feb 2024 14:57:35	S	Std
ICV14 - 1	ICV14	17605	4.1771	-	7470A	PPB	16 Feb 2024 15:00:22	U	SMPL
ICB14 - 1	ICB14	33	-0.0088	-	7470A	PPB	16 Feb 2024 15:02:37	U	SMPL
CCV15 - 1	CCV15	20276	4.8133	-	7470A	PPB	16 Feb 2024 15:04:55	U	SMPL
CCB15 - 1	CCB15	36	-0.0081	-	7470A	PPB	16 Feb 2024 15:07:10	U	SMPL
CRA - 1	CRA	953	0.2104	-	7470A	PPB	16 Feb 2024 15:09:28	U	SMPL
HighStd - 1	HighStd	41337	9.8303	-	7470A	PPB	16 Feb 2024 15:11:43	U	SMPL
ChkStd - 1	ChkStd	30045	7.1404	-	7470A	PPB	16 Feb 2024 15:13:58	U	SMPL
PB159046BL - 1	PBW	0	-0.0167	-	7470A	PPB	16 Feb 2024 15:16:17	U	SMPL
PB159046BS - 1	LCSW	16857	3.9989	-	7470A	PPB	16 Feb 2024 15:18:36	U	SMPL
P1461-03 - 1	TP-2	90	0.0048	-	7470A	PPB	16 Feb 2024 15:20:52	U	SMPL
P1461-03DUP - 1	TP-2DUP	79	0.0022	-	7470A	PPB	16 Feb 2024 15:26:53	U	SMPL
P1461-03MS - 1	TP-2MS	18355	4.3557	-	7470A	PPB	16 Feb 2024 15:29:10	U	SMPL
P1461-03MSD - 1	TP-2MSD	15963	3.7859	-	7470A	PPB	16 Feb 2024 15:31:26	U	SMPL
P1461-06 - 1	TP-1	140	0.0167	-	7470A	PPB	16 Feb 2024 15:33:44	U	SMPL
CCV16 - 1	CCV16	20025	4.7535	-	7470A	PPB	16 Feb 2024 15:36:02	U	SMPL
CCB16 - 1	CCB16	24	-0.0109	-	7470A	PPB	16 Feb 2024 15:38:17	U	SMPL
P1463-02 - 1	SP-F-COMP	289	0.0522	-	7470A	PPB	16 Feb 2024 15:40:35	U	SMPL
P1464-04 - 1	PN-1(SOUTH-AREA)	295	0.0536	-	7470A	PPB	16 Feb 2024 15:42:51	U	SMPL
P1464-05 - 1	PN-2(NORTH-AREA)	317	0.0589	-	7470A	PPB	16 Feb 2024 15:45:07	U	SMPL
P1464-06 - 1	PN-3(SMALL-STOCK-PILE)	325	0.0608	-	7470A	PPB	16 Feb 2024 15:47:22	U	SMPL
P1471-01 - 1	30992	867	0.1899	-	7470A	PPB	16 Feb 2024 15:49:38	U	SMPL
P1471-02 - 1	30984	701	0.1503	-	7470A	PPB	16 Feb 2024 15:51:54	U	SMPL
P1486-01 - 1	3381	286	0.0515	-	7470A	PPB	16 Feb 2024 15:54:11	U	SMPL
P1486-02 - 1	3420	272	0.0481	-	7470A	PPB	16 Feb 2024 15:56:28	U	SMPL
P1486-03 - 1	3418-3419	296	0.0539	-	7470A	PPB	16 Feb 2024 15:58:45	U	SMPL
P1488-01 - 1	2976	2062	0.4745	-	7470A	PPB	16 Feb 2024 16:01:00	U	SMPL
CCV17 - 1	CCV17	19166	4.5489	-	7470A	PPB	16 Feb 2024 16:03:15	U	SMPL
CCB17 - 1	CCB17	-45	-0.0274	-	7470A	PPB	16 Feb 2024 16:05:33	U	SMPL
PB159047BL - 1	PBW	188	0.0281	-	7470A	PPB	16 Feb 2024 16:11:46	U	SMPL
PB159047BS - 1	LCSW	17679	4.1947	-	7470A	PPB	16 Feb 2024 16:14:02	U	SMPL
P1471-04 - 1	30985	592	0.1244	-	7470A	PPB	16 Feb 2024 16:16:18	U	SMPL
P1471-04DUP - 1	30985DUP	735	0.1584	-	7470A	PPB	16 Feb 2024 16:18:37	U	SMPL
P1471-04MS - 1	30985MS	17525	4.1580	-	7470A	PPB	16 Feb 2024 16:20:53	U	SMPL
P1471-04MSD - 1	30985MSD	15487	3.6725	-	7470A	PPB	16 Feb 2024 16:23:09	U	SMPL
P1484-01 - 1	382	150	0.0191	-	7470A	PPB	16 Feb 2024 16:25:28	U	SMPL
P1484-02 - 1	386	840	0.1834	-	7470A	PPB	16 Feb 2024 16:27:48	U	SMPL
P1486-04 - 1	3397	680	0.1453	-	7470A	PPB	16 Feb 2024 16:30:05	U	SMPL
P1486-05 - 1	3428	848	0.1854	-	7470A	PPB	16 Feb 2024 16:32:20	U	SMPL
CCV18 - 1	CCV18	19314	4.5842	-	7470A	PPB	16 Feb 2024 16:34:36	U	SMPL
CCB18 - 1	CCB18	154	0.0200	-	7470A	PPB	16 Feb 2024 16:36:52	U	SMPL
P1486-06 - 1	3417	833	0.1818	-	7470A	PPB	16 Feb 2024 16:39:10	U	SMPL
PB159009TB - 1		220	0.0358	-	7470A	PPB	16 Feb 2024 16:41:26	U	SMPL
PB159030TB - 1		143	0.0174	-	7470A	PPB	16 Feb 2024 16:46:19	U	SMPL
P1461-03LX5 - 1		43	-0.0064	-	7470A	PPB	16 Feb 2024 16:48:34	U	SMPL
P1461-03A - 1		17111	4.0594	-	7470A	PPB	16 Feb 2024 16:50:50	U	SMPL
P1471-04LX5 - 1		124	0.0129	-	7470A	PPB	16 Feb 2024 16:53:07	U	SMPL
P1471-04A - 1		16666	3.9534	-	7470A	PPB	16 Feb 2024 16:55:26	U	SMPL
CCV19 - 1	CCV19	19874	4.7176	-	7470A	PPB	16 Feb 2024 16:57:43	U	SMPL
CCB19 - 1	CCB19	-22	-0.0219	-	7470A	PPB	16 Feb 2024 17:00:02	U	SMPL