



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129577

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/16/2024	16:37
2	CAL2	1	Water	NA	NA	20.3	7.00	02/16/2024	16:39
3	CAL3	1	Water	NA	NA	20.3	10.02	02/16/2024	16:40
4	ICV	1	Water	NA	NA	20.4	7.00	02/16/2024	16:45
5	CCV1	1	Water	NA	NA	20.3	2.01	02/16/2024	16:46
6	P1495-01	1	Water	NA	NA	20.1	5.09	02/16/2024	16:47
7	P1516-02	1	Water	NA	NA	20.2	5.41	02/16/2024	16:48
8	P1516-02DUP	1	Water	NA	NA	20.3	5.42	02/16/2024	16:49
9	CCV2	1	Water	NA	NA	20.3	12.03	02/16/2024	16:50

48 AB 1205-96
1205-96

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH P1516 WorkList ID : 177969 Department : Wet-Chemistry Date : 02-16-2024 14:22:45

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1495-01	A HAW-RIVER-MAIN-CELL	Water	pH	Cool 4 deg C	STER05	E21	02/12/2024	9040C
P1516-02	A 00042	Water	pH	Cool 4 deg C	PSEG03	E21	02/16/2024	9040C

Date/Time	02-16-24	17:00
Raw Sample Received by:	CP SM	
Raw Sample Relinquished by:	Jd Cdc	