



## Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB129578

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W3070         |
| BUFFER PH 7.00 GREEN 1PINT PK6 | W2981         |
| PH 10.01 BUFFER,COLOR CD 475ML | W3043         |
| buffer solution pH 7 yellow    | W3071         |
| Buffer Solution, PH2 (500ml)   | W3005         |
| Buffer Solution, PH12 (500ml)  | W3072         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.2             | 4.01        | 02/16/2024 | 12:50     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.3             | 7.00        | 02/16/2024 | 12:51     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.3             | 10.02       | 02/16/2024 | 12:55     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.4             | 7.02        | 02/16/2024 | 13:00     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.5             | 2.01        | 02/16/2024 | 13:05     |
| 6   | P1514-02    | 1  | Solid  | 20.02       | 20          | 23.1             | 5.54        | 02/16/2024 | 13:10     |
| 7   | P1514-02DUP | 1  | Solid  | 20.03       | 20          | 23.2             | 5.55        | 02/16/2024 | 13:11     |
| 8   | P1516-01    | 1  | Solid  | 20.02       | 20          | 23.6             | 5.53        | 02/16/2024 | 13:15     |
| 9   | P1516-03    | 1  | Solid  | 20.01       | 20          | 23.6             | 5.62        | 02/16/2024 | 13:24     |
| 10  | CCV2        | 1  | Water  | NA          | NA          | 20.2             | 12.02       | 02/16/2024 | 13:25     |