


LB129582

7470A

INSTRUMENT ID: CV1

μ Abs.:

4604E



Linear

A= 0.0000e+000

B= 2.2215e-004

C= 8.1585e-004

Rho= 0.9988757

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	U/D
0.0	0.000	0.009	0.009	35	0.000	35	0				0/0D
0.2	0.200	0.236	0.036	1058	0.0 %	1058					18
2.5	2.500	2.606	0.106	11725	0.0 %	11725					4
5.0	5.000	4.962	-0.038	22332	0.0 %	22332					-1
7.5	7.500	7.158	-0.342	32219	0.0 %	32219					-5
10.0	10.000	10.230	0.230	46046	0.0 %	46046					2

LB129582 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Feb 2024 12:16:17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	35	-	0.0000	7470A	PPB	19 Feb 2024 12:23:21	S	Std
0.2 - 1	50.2	1058	-	0.2000	7470A	PPB	19 Feb 2024 12:27:47	S	Std
2.5 - 1	52.5	11725	-	2.5000	7470A	PPB	19 Feb 2024 12:30:04	S	Std
5.0 - 1	55	22332	-	5.0000	7470A	PPB	19 Feb 2024 12:32:21	S	Std
7.5 - 1	57.5	32219	-	7.5000	7470A	PPB	19 Feb 2024 12:34:39	S	Std
10.0 - 1	510	46046	-	10.0000	7470A	PPB	19 Feb 2024 12:44:18	S	Std
ICV14 - 1	ICV14	18665	4.1472	-	7470A	PPB	19 Feb 2024 12:50:51	U	SMPL
ICB14 - 1	ICB14	54	0.0128	-	7470A	PPB	19 Feb 2024 12:53:06	U	SMPL
CCV45 - 1	CCV45	22350	4.9659	-	7470A	PPB	19 Feb 2024 12:55:24	U	SMPL
CCB45 - 1	CCB45	146	0.0332	-	7470A	PPB	19 Feb 2024 12:57:39	U	SMPL
CRA - 1	CRA	909	0.2027	-	7470A	PPB	19 Feb 2024 12:59:56	U	SMPL
HighStd - 1	HighStd	47148	10.4747	-	7470A	PPB	19 Feb 2024 13:02:11	U	SMPL
ChkStd - 1	ChkStd	29781	6.6167	-	7470A	PPB	19 Feb 2024 13:07:13	U	SMPL
PB159075BL - 1	PBW	270	0.0608	-	7470A	PPB	19 Feb 2024 13:11:25	U	SMPL
PB159075BS - 1	LCSW	20315	4.5138	-	7470A	PPB	19 Feb 2024 13:13:41	U	SMPL
P1505-04 - 1	WC-1S	1594	0.3549	-	7470A	PPB	19 Feb 2024 13:15:57	U	SMPL
P1505-08 - 1	WC-1D	1795	0.3996	-	7470A	PPB	19 Feb 2024 13:18:16	U	SMPL
P1505-12 - 1	WC-2S	1824	0.4060	-	7470A	PPB	19 Feb 2024 13:20:33	U	SMPL
P1505-16 - 1	WC-2D	1871	0.4165	-	7470A	PPB	19 Feb 2024 13:22:50	U	SMPL
P1505-16DUP - 1	WC-2DDUP	1637	0.3645	-	7470A	PPB	19 Feb 2024 13:25:05	U	SMPL
CCV46 - 1	CCV46	22043	4.8977	-	7470A	PPB	19 Feb 2024 13:27:20	U	SMPL
CCB46 - 1	CCB46	-7	-0.0007	-	7470A	PPB	19 Feb 2024 13:29:35	U	SMPL
P1505-16MS - 1	WC-2DMS	21240	4.7193	-	7470A	PPB	19 Feb 2024 13:31:54	U	SMPL
P1505-16MSD - 1	WC-2DMSD	17531	3.8953	-	7470A	PPB	19 Feb 2024 13:34:10	U	SMPL
PB159045TB - 1	PB159045TB	62	0.0146	-	7470A	PPB	19 Feb 2024 13:40:07	U	SMPL
P1505-16LX5 - 1		457	0.1023	-	7470A	PPB	19 Feb 2024 13:42:22	U	SMPL
P1505-16A - 1		20165	4.4805	-	7470A	PPB	19 Feb 2024 13:44:38	U	SMPL
CCV47 - 1	CCV47	21650	4.8104	-	7470A	PPB	19 Feb 2024 13:46:55	U	SMPL
CCB47 - 1	CCB47	141	0.0321	-	7470A	PPB	19 Feb 2024 13:49:14	U	SMPL