



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 2/20/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 16:30
In Date: 02/19/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 02/20/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB129586

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P1519-05	SVOC-GPC-BLANK	1	1.00	1.00	2.00	2.00	100.0	
P1519-06	PEST-GPC-BLANK	2	1.00	1.00	2.00	2.00	100.0	
P1519-07	PEST-GPC-BLANK-SPIKE	3	1.00	1.00	2.00	2.00	100.0	
P1519-08	PCB-GPC-BLANK	4	1.00	1.00	2.00	2.00	100.0	
P1519-09	PCB-GPC-BLANK-SPIKE	5	1.00	1.00	2.00	2.00	100.0	
P1519-10	SVOC-GPC2-BLANK	6	1.00	1.00	2.00	2.00	100.0	
P1519-11	PEST-GPC2-BLANK	7	1.00	1.00	2.00	2.00	100.0	
P1519-12	PEST-GPC2-BLANK-SPIKE	8	1.00	1.00	2.00	2.00	100.0	
P1519-13	PCB-GPC-BLANK-SPIKE	9	1.00	1.00	2.00	2.00	100.0	
P1519-14	PCB-GCP2-BLANK-SPIKE	10	1.00	1.00	2.00	2.00	100.0	
P1523-01	20240763-COMPOSITE-39N-N-3-01	11	1.15	8.81	9.96	4.3	35.8	sludge sample
P1523-02	20240764-COMPOSITE-39N-N-3-02	12	1.16	8.80	9.96	4.15	34.0	sludge sample
P1523-03	20240765-COMPOSITE-39N-N-3-03	13	1.15	8.83	9.98	3.95	31.7	sludge sample
P1523-04	P1523-03MS	14	1.15	8.83	9.98	3.95	31.7	sludge sample
P1523-05	P1523-03MSD	15	1.15	8.83	9.98	3.95	31.7	sludge sample
P1523-06	20240766-AMENDED-COMP-39N-N-3-01	16	1.15	8.79	9.94	4.8	41.5	
P1523-07	20240767-AMENDED-COMP-39N-N-3-02	17	1.16	8.47	9.63	4.57	40.3	
P1523-08	20240768-AMENDED-COMP-39N-N-3-03	18	1.13	8.71	9.84	4.58	39.6	
P1523-09	P1523-08MS	19	1.13	8.71	9.84	4.58	39.6	
P1523-10	P1523-08MSD	20	1.13	8.71	9.84	4.58	39.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$